

Report of Analysis

Client:	PSEG	Date Collected:	10/01/25
Project:	Moorestown HQ	Date Received:	10/01/25
Client Sample ID:	MOO-25-0282RE	SDG No.:	Q3262
Lab Sample ID:	Q3262-04RE	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	59.6
Sample Wt/Vol:	4.7 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VW032343.D	1	10/03/25 11:10	VW100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	0.0089	U	0.0020	0.0089	mg/Kg
74-87-3	Chloromethane	0.0089	U	0.0020	0.0089	mg/Kg
75-01-4	Vinyl Chloride	0.0089	U	0.0014	0.0089	mg/Kg
74-83-9	Bromomethane	0.0089	U	0.0019	0.0089	mg/Kg
75-00-3	Chloroethane	0.0089	U	0.0022	0.0089	mg/Kg
75-69-4	Trichlorofluoromethane	0.0089	U	0.0022	0.0089	mg/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	0.0089	U	0.0019	0.0089	mg/Kg
75-35-4	1,1-Dichloroethene	0.0089	U	0.0018	0.0089	mg/Kg
67-64-1	Acetone	0.045	U	0.0085	0.045	mg/Kg
75-15-0	Carbon Disulfide	0.0089	U	0.0019	0.0089	mg/Kg
1634-04-4	Methyl tert-butyl Ether	0.0089	U	0.0013	0.0089	mg/Kg
79-20-9	Methyl Acetate	0.0089	U	0.0027	0.0089	mg/Kg
75-09-2	Methylene Chloride	0.018	U	0.0063	0.018	mg/Kg
156-60-5	trans-1,2-Dichloroethene	0.0089	U	0.0015	0.0089	mg/Kg
75-34-3	1,1-Dichloroethane	0.0089	U	0.0014	0.0089	mg/Kg
110-82-7	Cyclohexane	0.0089	U	0.0014	0.0089	mg/Kg
78-93-3	2-Butanone	0.045	U	0.012	0.045	mg/Kg
56-23-5	Carbon Tetrachloride	0.0089	U	0.0017	0.0089	mg/Kg
156-59-2	cis-1,2-Dichloroethene	0.0089	U	0.0013	0.0089	mg/Kg
74-97-5	Bromochloromethane	0.0089	U	0.0021	0.0089	mg/Kg
67-66-3	Chloroform	0.0089	U	0.0015	0.0089	mg/Kg
71-55-6	1,1,1-Trichloroethane	0.0089	U	0.0017	0.0089	mg/Kg
108-87-2	Methylcyclohexane	0.0089	U	0.0016	0.0089	mg/Kg
71-43-2	Benzene	0.0089	U	0.0014	0.0089	mg/Kg
107-06-2	1,2-Dichloroethane	0.0089	U	0.0014	0.0089	mg/Kg
79-01-6	Trichloroethene	0.0089	U	0.0014	0.0089	mg/Kg
78-87-5	1,2-Dichloropropane	0.0089	U	0.0016	0.0089	mg/Kg
75-27-4	Bromodichloromethane	0.0089	U	0.0014	0.0089	mg/Kg
108-10-1	4-Methyl-2-Pentanone	0.045	U	0.0064	0.045	mg/Kg
108-88-3	Toluene	0.0032	J	0.0014	0.0089	mg/Kg

Report of Analysis

Client:	PSEG	Date Collected:	10/01/25
Project:	Moorestown HQ	Date Received:	10/01/25
Client Sample ID:	MOO-25-0282RE	SDG No.:	Q3262
Lab Sample ID:	Q3262-04RE	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	59.6
Sample Wt/Vol:	4.7 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VW032343.D	1	10/03/25 11:10	VW100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.0089	U	0.0012	0.0089	mg/Kg
10061-01-5	cis-1,3-Dichloropropene	0.0089	U	0.0011	0.0089	mg/Kg
79-00-5	1,1,2-Trichloroethane	0.0089	U	0.0016	0.0089	mg/Kg
591-78-6	2-Hexanone	0.045	U	0.0066	0.045	mg/Kg
124-48-1	Dibromochloromethane	0.0089	U	0.0016	0.0089	mg/Kg
106-93-4	1,2-Dibromoethane	0.0089	U	0.0016	0.0089	mg/Kg
127-18-4	Tetrachloroethene	0.0089	U	0.0019	0.0089	mg/Kg
108-90-7	Chlorobenzene	0.0089	U	0.0016	0.0089	mg/Kg
100-41-4	Ethyl Benzene	0.0089	U	0.0012	0.0089	mg/Kg
179601-23-1	m/p-Xylenes	0.0059	J	0.0022	0.018	mg/Kg
95-47-6	o-Xylene	0.0025	J	0.0015	0.0089	mg/Kg
100-42-5	Styrene	0.0089	U	0.0013	0.0089	mg/Kg
75-25-2	Bromoform	0.0089	U	0.0015	0.0089	mg/Kg
98-82-8	Isopropylbenzene	0.0089	U	0.0014	0.0089	mg/Kg
79-34-5	1,1,2,2-Tetrachloroethane	0.0089	U	0.0022	0.0089	mg/Kg
541-73-1	1,3-Dichlorobenzene	0.0089	U	0.0031	0.0089	mg/Kg
106-46-7	1,4-Dichlorobenzene	0.0089	U	0.0028	0.0089	mg/Kg
95-50-1	1,2-Dichlorobenzene	0.0089	U	0.0026	0.0089	mg/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	0.0089	U	0.0033	0.0089	mg/Kg
120-82-1	1,2,4-Trichlorobenzene	0.0089	U	0.0053	0.0089	mg/Kg
87-61-6	1,2,3-Trichlorobenzene	0.0089	U	0.0057	0.0089	mg/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	68.3		63 - 155	137%	SPK: 50
1868-53-7	Dibromofluoromethane	57.2		70 - 134	114%	SPK: 50
2037-26-5	Toluene-d8	39.9		74 - 123	80%	SPK: 50
460-00-4	4-Bromofluorobenzene	28.3		17 - 146	57%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	173000	7.959			
540-36-3	1,4-Difluorobenzene	383000	8.849			
3114-55-4	Chlorobenzene-d5	288000	11.629			
3855-82-1	1,4-Dichlorobenzene-d4	68400	13.556			

Report of Analysis

Client:	PSEG	Date Collected:	10/01/25
Project:	Moorestown HQ	Date Received:	10/01/25
Client Sample ID:	MOO-25-0282RE	SDG No.:	Q3262
Lab Sample ID:	Q3262-04RE	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	59.6
Sample Wt/Vol:	4.7 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VW032343.D	1	10/03/25 11:10	VW100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products