



284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

CASE NARRATIVE

CDM Smith

Project Name: MWCO CJO WTP Edison 295110

Project # N/A

Order ID # Q3266

**Test Name: VOC-TCLVOA-10,SVOC-TCL BNA -
20,EPH,Herbicide,PCB,Pesticide-TCL,Mercury,Metals ICP-TAL**

A. Number of Samples and Date of Receipt:

2 Solid samples were received on 10/01/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: VOC-TCLVOA-10,SVOC-TCL BNA -20,EPH,Herbicide,PCB,Pesticide-TCL,Mercury,Metals ICP-TAL. This data package contains results for VOC-TCLVOA-10(8260D), SVOC-TCL BNA -20(8270E),EPH(NJEPH),Herbicide(8151A),PCB(8082A),Pesticide-TCL(8081B),Mercury(7471B),Metals ICP-TAL(6010D).

C. Analytical Techniques:

VOC-TCLVOA-10 : The analysis performed on instrument MSVOA_W were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868.The analysis of VOC-TCLVOA-10 was based on method 8260D.

SVOC-TCL BNA -20 : The samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df. The samples were analyzed on instrument BNA_P using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGA. The analysis of SVOC-TCL BNA -20 was based on method 8270E and extraction was done based on method 3541.

Pesticide-TCL : The analysis was performed on instrument ECD_L. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalog # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 7HMG017- 11.The analysis of Pesticide-TCLs was based on method 8081B and extraction was done based on method 3541.



284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

PCB : The analyses were performed on instrument GCECD_P. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 μ m df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 μ m; Catalogue # 7HM-G017-11. The analyses were performed on instrument GCECD_O. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 μ m df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 μ m; Catalogue # 7HM-G017-11. The analysis of PCBs was based on method 8082A and extraction was done based on method 3541.

Herbicide : The analysis was performed on instrument ECD_S. The front column is RTX-CLPesticides which is 30 meters, 0.32 mm ID, 0.5 μ m df,; Catalog # 11139. The rear column is RTX-CLPesticides2 which is 30 meters, 0.32 mm ID, 0.25 μ m df, Catalog #: 11324. The analysis of Herbicides was based on method 8151A and extraction was done based on method 3541.

EPH : The analyses were performed on instrument FID_D. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 μ m df, catalog 10224. The analysis were performed on instrument FID_E. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 μ m df, catalog 10224. The analysis of EPHs was based on method NJEPH and extraction was done based on method 3541.

Mercury, Metals ICP-TAL : The analysis of Metals ICP-TAL was based on method 6010D, digestion based on method 3050 (soils). The analysis and digestion of Mercury was based on method 7471B.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis except following

SVOC-TCL BNA -20 : ENV-3-6FT [Terphenyl-d14 - 116%], As per SOW one base surrogate allowed fail, Therefore no further corrective action was taken. ENV-3-6FTMS [Nitrobenzene-d5 - 120%, Phenol-d6 - 115%, Terphenyl-d14 - 139%] and ENV-3-6FTMSD [Nitrobenzene-d5 - 112%, Terphenyl-d14 - 127%]. Failed same surrogate in MS/MSD due to matrix interference, Therefore no further corrective action was taken.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS recoveries met the requirements for all compounds except following

SVOC-TCL BNA -20 : The MS {Q3266-01MS} with File ID: BF143864.D recoveries met the requirements for all compounds except for 2,6-Dinitrotoluene[132%], Recovery failed due to matrix interference, Therefore no further corrective action was taken.



284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

Pesticide-TCL : The MS {Q3266-01MS} with File ID: PL097503.D recoveries met the requirements for all compounds except for [Heptachlor(2)- 60%], [Methoxychlor(2)- 53%] due to matrix interference.

Herbicide : The MS {Q3266-01MS} with File ID: PS031959.D recoveries met the requirements for all compounds except for [Dinoseb(1)0% - Dinoseb(2)0%] due to matrix interference.

Mercury, Metals ICP-TAL : The Matrix Spike (ENV-4-6FTMS) analysis met criteria for all compounds except for Antimony, Chromium, Copper, Potassium and Vanadium due to Chemical Interference during Digestion Process.

The MSD recoveries met the requirements for all compounds except following
Pesticide-TCL : The MSD {Q3266-01MSD} with File ID: PL097504.D recoveries met the requirements for all compounds except for [Heptachlor(2)- 61%] due to matrix interference.

Herbicide : The MSD {Q3266-01MSD} with File ID: PS031960.D recoveries met the requirements for all compounds except for [Dinoseb(1)0% - Dinoseb(2)0%] due to matrix interference.

Mercury, Metals ICP-TAL : The Matrix Spike Duplicate (ENV-4-6FTMSD) analysis met criteria for all compounds except for Antimony, Cobalt, Copper, Nickel and Potassium due to Chemical Interference during Digestion process.

The RPD recoveries met criteria.

The Blank Spike met requirements for all compounds except following
VOC-TCLVOA-10 : The Blank Spike for {VW1002SBS01} with File ID: VW032324.D met requirements for all compounds except for Methylene Chloride[134%]. Failing high but associated samples have no positive hit for this compound therefore no corrective action was taken.

The Blank Spike Duplicate met requirements for all compounds

The Blank Spike met requirements for all compounds except following
SVOC-TCL BNA -20 : The Blank Spike for {PB169989BS} with File ID: BP025926.D met requirements for all compounds except for 1,2,4,5-Tetrachlorobenzene[106%]. there is no positive hit for failed compound in associated samples, Therefore no further corrective action was taken.

The Blank Spike Duplicate met requirements for all compounds

The Blank analysis did not indicate the presence of lab contamination.



284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

The Initial Calibration met the requirements except following
VOC-TCLVOA-10 : The %RSD is greater than 20% in the Initial Calibration method (82W100125S.M) for Methylene chloride, this compound is passing on Linear regression.

SVOC-TCL BNA -20 : The %RSD is greater than 20% in the Method 8270-BF092325.M for 2-Nitrophenol & 2,4-Dinitrotoluene passing on Linear regression, and 2,4-Dinitrophenol, 4,6-Dinitro-2-methylphenol, they are passing on Quadratic regression.

The %RSD is greater than 20% in the Initial Calibration (Method 8270-BF100625.M) for 2,4-Dinitrophenol this Compounds is passing on Linear regression.

The %RSD is greater than 20% in the Initial Calibration (Method 8270-BP091225.M) for 2,4-Dinitrophenol & 4-Nitrophenol. These Compounds are passing on Linear regression.

The Continuous Calibration met the requirements except following
SVOC-TCL BNA -20 : The Continuous Calibration File ID BF143853.D met the requirements except for 2,3,4,6-Tetrachlorophenol,2,4-Dinitrophenol,2,6-Dinitrotoluene,4-Bromophenyl-phenylether,Hexachlorobenzene,Pentachlorophenol and 2,4,6-Tribromophenol. There is no positive hit in associated samples, therefor no further corrective action was taken.

The Continuous Calibration File ID BP025924.D met the requirements except for Benzaldehyde. There is no positive hit in associated samples, therefor no further corrective action was taken

PCB : The Continuous Calibration File ID PO114211.D met the requirements except for Aroclor-1260(Peak-04),Aroclor-1260(Peak-05) is failing in 1st column, however it is passed in 2nd column therefore no corrective action was taken.

The Tuning criteria met requirements.

The Duplicate analysis met criteria for all samples.

The Serial Dilution met criteria for all compounds except following
Mercury, Metals ICP-TAL : The Serial Dilution (ENV-4-6FTL) met criteria for all compounds except for Aluminum, Calcium, Chromium, Copper, Magnesium, Manganese and Zinc due to unknow sample matrix interference.



284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

E. Additional Comments:

The soil samples results are based on a dry weight basis.

SEMI-VOA : The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

Mercury, Metals ICP-TAL : The Post Digest Spike (ENV-4-6FTA) analysis met criteria for all compounds except for Potassium due to unknown chemical interference of matrix with the addition of spike amount after digestion and before analysis; matrix has suppression effect during addition of spike.

VOC-TCLVOA-10 : Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.
Trip Blank was not provided with this set of samples.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____