



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3266

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB56	Mercury	0.076	+/-0.2	U	0.20	CV	10/03/2025	08:45	LB137397

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3266

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB75	Mercury	0.076	+/-0.2	U	0.20	CV	10/03/2025	08:50	LB137397
CCB76	Mercury	0.076	+/-0.2	U	0.20	CV	10/03/2025	09:23	LB137397
CCB77	Mercury	0.076	+/-0.2	U	0.20	CV	10/03/2025	09:41	LB137397

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3266

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.6	+/-50	J	100	P	10/03/2025	13:12	LB137426
	Antimony	6.76	+/-25	U	50.0	P	10/03/2025	13:12	LB137426
	Arsenic	5.12	+/-10	U	20.0	P	10/03/2025	13:12	LB137426
	Barium	14.6	+/-50	U	100	P	10/03/2025	13:12	LB137426
	Beryllium	0.56	+/-3	U	6.00	P	10/03/2025	13:12	LB137426
	Cadmium	0.50	+/-3	U	6.00	P	10/03/2025	13:12	LB137426
	Calcium	234	+/-1000	U	2000	P	10/03/2025	13:12	LB137426
	Chromium	2.12	+/-5	U	10.0	P	10/03/2025	13:12	LB137426
	Cobalt	2.26	+/-15	U	30.0	P	10/03/2025	13:12	LB137426
	Copper	4.60	+/-10	U	20.0	P	10/03/2025	13:12	LB137426
	Iron	23.4	+/-50	U	100	P	10/03/2025	13:12	LB137426
	Lead	2.30	+/-6	U	12.0	P	10/03/2025	13:12	LB137426
	Magnesium	244	+/-1000	U	2000	P	10/03/2025	13:12	LB137426
	Manganese	5.94	+/-10	U	20.0	P	10/03/2025	13:12	LB137426
	Nickel	3.06	+/-20	U	40.0	P	10/03/2025	13:12	LB137426
	Potassium	918	+/-1000	U	2000	P	10/03/2025	13:12	LB137426
	Selenium	9.64	+/-10	U	20.0	P	10/03/2025	13:12	LB137426
	Silver	1.62	+/-5	U	10.0	P	10/03/2025	13:12	LB137426
	Sodium	868	+/-1000	U	2000	P	10/03/2025	13:12	LB137426
	Thallium	4.38	+/-20	U	40.0	P	10/03/2025	13:12	LB137426
	Vanadium	6.26	+/-20	U	40.0	P	10/03/2025	13:12	LB137426
	Zinc	16.7	+/-20	U	40.0	P	10/03/2025	13:12	LB137426

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3266

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	10/03/2025	13:41	LB137426
	Antimony	6.76	+/-25	U	50.0	P	10/03/2025	13:41	LB137426
	Arsenic	5.12	+/-10	U	20.0	P	10/03/2025	13:41	LB137426
	Barium	14.6	+/-50	U	100	P	10/03/2025	13:41	LB137426
	Beryllium	0.56	+/-3	U	6.00	P	10/03/2025	13:41	LB137426
	Cadmium	0.50	+/-3	U	6.00	P	10/03/2025	13:41	LB137426
	Calcium	234	+/-1000	U	2000	P	10/03/2025	13:41	LB137426
	Chromium	2.12	+/-5	U	10.0	P	10/03/2025	13:41	LB137426
	Cobalt	2.26	+/-15	U	30.0	P	10/03/2025	13:41	LB137426
	Copper	4.60	+/-10	U	20.0	P	10/03/2025	13:41	LB137426
	Iron	23.4	+/-50	U	100	P	10/03/2025	13:41	LB137426
	Lead	2.30	+/-6	U	12.0	P	10/03/2025	13:41	LB137426
	Magnesium	244	+/-1000	U	2000	P	10/03/2025	13:41	LB137426
	Manganese	5.94	+/-10	U	20.0	P	10/03/2025	13:41	LB137426
	Nickel	3.06	+/-20	U	40.0	P	10/03/2025	13:41	LB137426
	Potassium	918	+/-1000	U	2000	P	10/03/2025	13:41	LB137426
	Selenium	9.64	+/-10	U	20.0	P	10/03/2025	13:41	LB137426
	Silver	1.62	+/-5	U	10.0	P	10/03/2025	13:41	LB137426
	Sodium	868	+/-1000	U	2000	P	10/03/2025	13:41	LB137426
	Thallium	4.38	+/-20	U	40.0	P	10/03/2025	13:41	LB137426
	Vanadium	6.26	+/-20	U	40.0	P	10/03/2025	13:41	LB137426
	Zinc	16.7	+/-20	U	40.0	P	10/03/2025	13:41	LB137426
CCB02	Aluminum	11.3	+/-50	U	100	P	10/03/2025	15:04	LB137426
	Antimony	6.76	+/-25	U	50.0	P	10/03/2025	15:04	LB137426
	Arsenic	5.12	+/-10	U	20.0	P	10/03/2025	15:04	LB137426
	Barium	14.6	+/-50	U	100	P	10/03/2025	15:04	LB137426
	Beryllium	0.56	+/-3	U	6.00	P	10/03/2025	15:04	LB137426
	Cadmium	0.50	+/-3	U	6.00	P	10/03/2025	15:04	LB137426
	Calcium	234	+/-1000	U	2000	P	10/03/2025	15:04	LB137426
	Chromium	2.12	+/-5	U	10.0	P	10/03/2025	15:04	LB137426
	Cobalt	2.26	+/-15	U	30.0	P	10/03/2025	15:04	LB137426
	Copper	4.60	+/-10	U	20.0	P	10/03/2025	15:04	LB137426
	Iron	23.4	+/-50	U	100	P	10/03/2025	15:04	LB137426
	Lead	2.30	+/-6	U	12.0	P	10/03/2025	15:04	LB137426
	Magnesium	244	+/-1000	U	2000	P	10/03/2025	15:04	LB137426
	Manganese	5.94	+/-10	U	20.0	P	10/03/2025	15:04	LB137426
	Nickel	3.06	+/-20	U	40.0	P	10/03/2025	15:04	LB137426
	Potassium	918	+/-1000	U	2000	P	10/03/2025	15:04	LB137426
	Selenium	9.64	+/-10	U	20.0	P	10/03/2025	15:04	LB137426

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3266

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	10/03/2025	15:04	LB137426
	Sodium	868	+/-1000	U	2000	P	10/03/2025	15:04	LB137426
	Thallium	10.1	+/-20	J	40.0	P	10/03/2025	15:04	LB137426
	Vanadium	6.26	+/-20	U	40.0	P	10/03/2025	15:04	LB137426
	Zinc	16.7	+/-20	U	40.0	P	10/03/2025	15:04	LB137426
CCB03	Aluminum	11.3	+/-50	U	100	P	10/03/2025	16:00	LB137426
	Antimony	6.76	+/-25	U	50.0	P	10/03/2025	16:00	LB137426
	Arsenic	5.12	+/-10	U	20.0	P	10/03/2025	16:00	LB137426
	Barium	14.6	+/-50	U	100	P	10/03/2025	16:00	LB137426
	Beryllium	0.56	+/-3	U	6.00	P	10/03/2025	16:00	LB137426
	Cadmium	0.50	+/-3	U	6.00	P	10/03/2025	16:00	LB137426
	Calcium	234	+/-1000	U	2000	P	10/03/2025	16:00	LB137426
	Chromium	2.12	+/-5	U	10.0	P	10/03/2025	16:00	LB137426
	Cobalt	2.26	+/-15	U	30.0	P	10/03/2025	16:00	LB137426
	Copper	4.60	+/-10	U	20.0	P	10/03/2025	16:00	LB137426
	Iron	23.4	+/-50	U	100	P	10/03/2025	16:00	LB137426
	Lead	2.30	+/-6	U	12.0	P	10/03/2025	16:00	LB137426
	Magnesium	244	+/-1000	U	2000	P	10/03/2025	16:00	LB137426
	Manganese	5.94	+/-10	U	20.0	P	10/03/2025	16:00	LB137426
	Nickel	3.06	+/-20	U	40.0	P	10/03/2025	16:00	LB137426
	Potassium	918	+/-1000	U	2000	P	10/03/2025	16:00	LB137426
	Selenium	9.64	+/-10	U	20.0	P	10/03/2025	16:00	LB137426
	Silver	1.62	+/-5	U	10.0	P	10/03/2025	16:00	LB137426
	Sodium	868	+/-1000	U	2000	P	10/03/2025	16:00	LB137426
	Thallium	4.38	+/-20	U	40.0	P	10/03/2025	16:00	LB137426
	Vanadium	6.26	+/-20	U	40.0	P	10/03/2025	16:00	LB137426
	Zinc	16.7	+/-20	U	40.0	P	10/03/2025	16:00	LB137426
CCB04	Aluminum	11.3	+/-50	U	100	P	10/03/2025	16:49	LB137426
	Antimony	6.76	+/-25	U	50.0	P	10/03/2025	16:49	LB137426
	Arsenic	5.12	+/-10	U	20.0	P	10/03/2025	16:49	LB137426
	Barium	14.6	+/-50	U	100	P	10/03/2025	16:49	LB137426
	Beryllium	0.56	+/-3	U	6.00	P	10/03/2025	16:49	LB137426
	Cadmium	0.50	+/-3	U	6.00	P	10/03/2025	16:49	LB137426
	Calcium	234	+/-1000	U	2000	P	10/03/2025	16:49	LB137426
	Chromium	2.12	+/-5	U	10.0	P	10/03/2025	16:49	LB137426
	Cobalt	2.26	+/-15	U	30.0	P	10/03/2025	16:49	LB137426
	Copper	4.60	+/-10	U	20.0	P	10/03/2025	16:49	LB137426
	Iron	23.4	+/-50	U	100	P	10/03/2025	16:49	LB137426
	Lead	2.30	+/-6	U	12.0	P	10/03/2025	16:49	LB137426

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3266

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	10/03/2025	16:49	LB137426
	Manganese	5.94	+/-10	U	20.0	P	10/03/2025	16:49	LB137426
	Nickel	3.06	+/-20	U	40.0	P	10/03/2025	16:49	LB137426
	Potassium	918	+/-1000	U	2000	P	10/03/2025	16:49	LB137426
	Selenium	9.64	+/-10	U	20.0	P	10/03/2025	16:49	LB137426
	Silver	1.62	+/-5	U	10.0	P	10/03/2025	16:49	LB137426
	Sodium	868	+/-1000	U	2000	P	10/03/2025	16:49	LB137426
	Thallium	4.38	+/-20	U	40.0	P	10/03/2025	16:49	LB137426
	Vanadium	6.26	+/-20	U	40.0	P	10/03/2025	16:49	LB137426
	Zinc	16.7	+/-20	U	40.0	P	10/03/2025	16:49	LB137426

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3266

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	10/07/2025	12:06	LB137447
	Antimony	6.76	+/-25	U	50.0	P	10/07/2025	12:06	LB137447
	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	12:06	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	12:06	LB137447
	Beryllium	0.56	+/-3	U	6.00	P	10/07/2025	12:06	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	12:06	LB137447
	Calcium	234	+/-1000	U	2000	P	10/07/2025	12:06	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	12:06	LB137447
	Cobalt	2.26	+/-15	U	30.0	P	10/07/2025	12:06	LB137447
	Copper	4.60	+/-10	U	20.0	P	10/07/2025	12:06	LB137447
	Iron	23.4	+/-50	U	100	P	10/07/2025	12:06	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	12:06	LB137447
	Magnesium	244	+/-1000	U	2000	P	10/07/2025	12:06	LB137447
	Manganese	5.94	+/-10	U	20.0	P	10/07/2025	12:06	LB137447
	Nickel	3.06	+/-20	U	40.0	P	10/07/2025	12:06	LB137447
	Potassium	918	+/-1000	U	2000	P	10/07/2025	12:06	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	12:06	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	12:06	LB137447
	Sodium	868	+/-1000	U	2000	P	10/07/2025	12:06	LB137447
	Thallium	5.24	+/-20	J	40.0	P	10/07/2025	12:06	LB137447
	Vanadium	6.26	+/-20	U	40.0	P	10/07/2025	12:06	LB137447
	Zinc	16.7	+/-20	U	40.0	P	10/07/2025	12:06	LB137447

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3266

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	10/07/2025	12:58	LB137447
	Antimony	6.76	+/-25	U	50.0	P	10/07/2025	12:58	LB137447
	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	12:58	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	12:58	LB137447
	Beryllium	0.56	+/-3	U	6.00	P	10/07/2025	12:58	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	12:58	LB137447
	Calcium	234	+/-1000	U	2000	P	10/07/2025	12:58	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	12:58	LB137447
	Cobalt	2.26	+/-15	U	30.0	P	10/07/2025	12:58	LB137447
	Copper	4.60	+/-10	U	20.0	P	10/07/2025	12:58	LB137447
	Iron	23.4	+/-50	U	100	P	10/07/2025	12:58	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	12:58	LB137447
	Magnesium	244	+/-1000	U	2000	P	10/07/2025	12:58	LB137447
	Manganese	5.94	+/-10	U	20.0	P	10/07/2025	12:58	LB137447
	Nickel	3.06	+/-20	U	40.0	P	10/07/2025	12:58	LB137447
	Potassium	918	+/-1000	U	2000	P	10/07/2025	12:58	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	12:58	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	12:58	LB137447
	Sodium	868	+/-1000	U	2000	P	10/07/2025	12:58	LB137447
	Thallium	4.38	+/-20	U	40.0	P	10/07/2025	12:58	LB137447
	Vanadium	6.26	+/-20	U	40.0	P	10/07/2025	12:58	LB137447
	Zinc	16.7	+/-20	U	40.0	P	10/07/2025	12:58	LB137447
CCB02	Aluminum	14.4	+/-50	J	100	P	10/07/2025	14:24	LB137447
	Antimony	6.76	+/-25	U	50.0	P	10/07/2025	14:24	LB137447
	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	14:24	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	14:24	LB137447
	Beryllium	0.56	+/-3	U	6.00	P	10/07/2025	14:24	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	14:24	LB137447
	Calcium	234	+/-1000	U	2000	P	10/07/2025	14:24	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	14:24	LB137447
	Cobalt	2.26	+/-15	U	30.0	P	10/07/2025	14:24	LB137447
	Copper	4.60	+/-10	U	20.0	P	10/07/2025	14:24	LB137447
	Iron	28.4	+/-50	J	100	P	10/07/2025	14:24	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	14:24	LB137447
	Magnesium	244	+/-1000	U	2000	P	10/07/2025	14:24	LB137447
	Manganese	5.94	+/-10	U	20.0	P	10/07/2025	14:24	LB137447
	Nickel	3.06	+/-20	U	40.0	P	10/07/2025	14:24	LB137447
	Potassium	918	+/-1000	U	2000	P	10/07/2025	14:24	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	14:24	LB137447

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3266

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	10/07/2025	14:24	LB137447
	Sodium	868	+/-1000	U	2000	P	10/07/2025	14:24	LB137447
	Thallium	5.15	+/-20	J	40.0	P	10/07/2025	14:24	LB137447
	Vanadium	6.26	+/-20	U	40.0	P	10/07/2025	14:24	LB137447
	Zinc	16.7	+/-20	U	40.0	P	10/07/2025	14:24	LB137447
CCB03	Aluminum	11.3	+/-50	U	100	P	10/07/2025	15:12	LB137447
	Antimony	6.76	+/-25	U	50.0	P	10/07/2025	15:12	LB137447
	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	15:12	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	15:12	LB137447
	Beryllium	0.56	+/-3	U	6.00	P	10/07/2025	15:12	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	15:12	LB137447
	Calcium	234	+/-1000	U	2000	P	10/07/2025	15:12	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	15:12	LB137447
	Cobalt	2.26	+/-15	U	30.0	P	10/07/2025	15:12	LB137447
	Copper	4.60	+/-10	U	20.0	P	10/07/2025	15:12	LB137447
	Iron	23.4	+/-50	U	100	P	10/07/2025	15:12	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	15:12	LB137447
	Magnesium	244	+/-1000	U	2000	P	10/07/2025	15:12	LB137447
	Manganese	5.94	+/-10	U	20.0	P	10/07/2025	15:12	LB137447
	Nickel	3.06	+/-20	U	40.0	P	10/07/2025	15:12	LB137447
	Potassium	918	+/-1000	U	2000	P	10/07/2025	15:12	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	15:12	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	15:12	LB137447
	Sodium	868	+/-1000	U	2000	P	10/07/2025	15:12	LB137447
	Thallium	6.22	+/-20	J	40.0	P	10/07/2025	15:12	LB137447
	Vanadium	6.26	+/-20	U	40.0	P	10/07/2025	15:12	LB137447
	Zinc	16.7	+/-20	U	40.0	P	10/07/2025	15:12	LB137447
CCB04	Aluminum	11.3	+/-50	U	100	P	10/07/2025	16:04	LB137447
	Antimony	6.76	+/-25	U	50.0	P	10/07/2025	16:04	LB137447
	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	16:04	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	16:04	LB137447
	Beryllium	0.56	+/-3	U	6.00	P	10/07/2025	16:04	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	16:04	LB137447
	Calcium	234	+/-1000	U	2000	P	10/07/2025	16:04	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	16:04	LB137447
	Cobalt	2.26	+/-15	U	30.0	P	10/07/2025	16:04	LB137447
	Copper	4.60	+/-10	U	20.0	P	10/07/2025	16:04	LB137447
	Iron	23.4	+/-50	U	100	P	10/07/2025	16:04	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	16:04	LB137447

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3266

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	10/07/2025	16:04	LB137447
	Manganese	5.94	+/-10	U	20.0	P	10/07/2025	16:04	LB137447
	Nickel	3.06	+/-20	U	40.0	P	10/07/2025	16:04	LB137447
	Potassium	918	+/-1000	U	2000	P	10/07/2025	16:04	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	16:04	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	16:04	LB137447
	Sodium	868	+/-1000	U	2000	P	10/07/2025	16:04	LB137447
	Thallium	5.51	+/-20	J	40.0	P	10/07/2025	16:04	LB137447
	Vanadium	6.26	+/-20	U	40.0	P	10/07/2025	16:04	LB137447
	Zinc	16.7	+/-20	U	40.0	P	10/07/2025	16:04	LB137447
CCB05	Aluminum	11.3	+/-50	U	100	P	10/07/2025	17:01	LB137447
	Antimony	6.76	+/-25	U	50.0	P	10/07/2025	17:01	LB137447
	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	17:01	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	17:01	LB137447
	Beryllium	0.56	+/-3	U	6.00	P	10/07/2025	17:01	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	17:01	LB137447
	Calcium	234	+/-1000	U	2000	P	10/07/2025	17:01	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	17:01	LB137447
	Cobalt	2.26	+/-15	U	30.0	P	10/07/2025	17:01	LB137447
	Copper	4.60	+/-10	U	20.0	P	10/07/2025	17:01	LB137447
	Iron	23.4	+/-50	U	100	P	10/07/2025	17:01	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	17:01	LB137447
	Magnesium	244	+/-1000	U	2000	P	10/07/2025	17:01	LB137447
	Manganese	5.94	+/-10	U	20.0	P	10/07/2025	17:01	LB137447
	Nickel	3.06	+/-20	U	40.0	P	10/07/2025	17:01	LB137447
	Potassium	918	+/-1000	U	2000	P	10/07/2025	17:01	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	17:01	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	17:01	LB137447
	Sodium	868	+/-1000	U	2000	P	10/07/2025	17:01	LB137447
	Thallium	5.07	+/-20	J	40.0	P	10/07/2025	17:01	LB137447
	Vanadium	6.26	+/-20	U	40.0	P	10/07/2025	17:01	LB137447
	Zinc	16.7	+/-20	U	40.0	P	10/07/2025	17:01	LB137447
CCB06	Aluminum	11.3	+/-50	U	100	P	10/07/2025	17:59	LB137447
	Antimony	6.76	+/-25	U	50.0	P	10/07/2025	17:59	LB137447
	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	17:59	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	17:59	LB137447
	Beryllium	0.56	+/-3	U	6.00	P	10/07/2025	17:59	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	17:59	LB137447
	Calcium	234	+/-1000	U	2000	P	10/07/2025	17:59	LB137447

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3266

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	17:59	LB137447
	Cobalt	2.26	+/-15	U	30.0	P	10/07/2025	17:59	LB137447
	Copper	4.60	+/-10	U	20.0	P	10/07/2025	17:59	LB137447
	Iron	23.4	+/-50	U	100	P	10/07/2025	17:59	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	17:59	LB137447
	Magnesium	244	+/-1000	U	2000	P	10/07/2025	17:59	LB137447
	Manganese	5.94	+/-10	U	20.0	P	10/07/2025	17:59	LB137447
	Nickel	3.06	+/-20	U	40.0	P	10/07/2025	17:59	LB137447
	Potassium	918	+/-1000	U	2000	P	10/07/2025	17:59	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	17:59	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	17:59	LB137447
	Sodium	868	+/-1000	U	2000	P	10/07/2025	17:59	LB137447
	Thallium	4.38	+/-20	U	40.0	P	10/07/2025	17:59	LB137447
	Vanadium	6.26	+/-20	U	40.0	P	10/07/2025	17:59	LB137447
	Zinc	16.7	+/-20	U	40.0	P	10/07/2025	17:59	LB137447
CCB07	Aluminum	11.3	+/-50	U	100	P	10/07/2025	18:48	LB137447
	Antimony	6.76	+/-25	U	50.0	P	10/07/2025	18:48	LB137447
	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	18:48	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	18:48	LB137447
	Beryllium	0.56	+/-3	U	6.00	P	10/07/2025	18:48	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	18:48	LB137447
	Calcium	234	+/-1000	U	2000	P	10/07/2025	18:48	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	18:48	LB137447
	Cobalt	2.26	+/-15	U	30.0	P	10/07/2025	18:48	LB137447
	Copper	4.60	+/-10	U	20.0	P	10/07/2025	18:48	LB137447
	Iron	23.4	+/-50	U	100	P	10/07/2025	18:48	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	18:48	LB137447
	Magnesium	244	+/-1000	U	2000	P	10/07/2025	18:48	LB137447
	Manganese	5.94	+/-10	U	20.0	P	10/07/2025	18:48	LB137447
	Nickel	3.06	+/-20	U	40.0	P	10/07/2025	18:48	LB137447
	Potassium	918	+/-1000	U	2000	P	10/07/2025	18:48	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	18:48	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	18:48	LB137447
	Sodium	868	+/-1000	U	2000	P	10/07/2025	18:48	LB137447
	Thallium	4.38	+/-20	U	40.0	P	10/07/2025	18:48	LB137447
	Vanadium	6.26	+/-20	U	40.0	P	10/07/2025	18:48	LB137447
	Zinc	16.7	+/-20	U	40.0	P	10/07/2025	18:48	LB137447
CCB08	Aluminum	11.3	+/-50	U	100	P	10/07/2025	19:24	LB137447
	Antimony	6.76	+/-25	U	50.0	P	10/07/2025	19:24	LB137447

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3266

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	19:24	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	19:24	LB137447
	Beryllium	0.56	+/-3	U	6.00	P	10/07/2025	19:24	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	19:24	LB137447
	Calcium	234	+/-1000	U	2000	P	10/07/2025	19:24	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	19:24	LB137447
	Cobalt	2.26	+/-15	U	30.0	P	10/07/2025	19:24	LB137447
	Copper	4.60	+/-10	U	20.0	P	10/07/2025	19:24	LB137447
	Iron	23.4	+/-50	U	100	P	10/07/2025	19:24	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	19:24	LB137447
	Magnesium	244	+/-1000	U	2000	P	10/07/2025	19:24	LB137447
	Manganese	5.94	+/-10	U	20.0	P	10/07/2025	19:24	LB137447
	Nickel	3.06	+/-20	U	40.0	P	10/07/2025	19:24	LB137447
	Potassium	918	+/-1000	U	2000	P	10/07/2025	19:24	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	19:24	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	19:24	LB137447
	Sodium	868	+/-1000	U	2000	P	10/07/2025	19:24	LB137447
	Thallium	4.38	+/-20	U	40.0	P	10/07/2025	19:24	LB137447
	Vanadium	6.26	+/-20	U	40.0	P	10/07/2025	19:24	LB137447
	Zinc	16.7	+/-20	U	40.0	P	10/07/2025	19:24	LB137447