

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>CDM Smith</u>	SDG No.: <u>Q3266</u>
Contract: <u>CAMP02</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	245000	255000	96	216000	294000	10/03/2025	13:20	LB137426
	Antimony	3.52			-50	50	10/03/2025	13:20	LB137426
	Arsenic	8.41			-20	20	10/03/2025	13:20	LB137426
	Barium	1.66	6.0	28	-94	106	10/03/2025	13:20	LB137426
	Beryllium	1.42			-6	6	10/03/2025	13:20	LB137426
	Cadmium	0.91	1.0	91	-5	7	10/03/2025	13:20	LB137426
	Calcium	236000	245000	96	208000	282000	10/03/2025	13:20	LB137426
	Chromium	46.6	52.0	90	42	62	10/03/2025	13:20	LB137426
	Cobalt	0.68			-30	30	10/03/2025	13:20	LB137426
	Copper	3.06	2.0	153	-18	22	10/03/2025	13:20	LB137426
	Iron	99600	101000	99	85600	116500	10/03/2025	13:20	LB137426
	Lead	0.97			-12	12	10/03/2025	13:20	LB137426
	Magnesium	247000	255000	97	216000	294000	10/03/2025	13:20	LB137426
	Manganese	-1.76	7.0	25	-13	27	10/03/2025	13:20	LB137426
	Nickel	3.93	2.0	196	-38	42	10/03/2025	13:20	LB137426
	Potassium	187			0	0	10/03/2025	13:20	LB137426
	Selenium	9.27			-20	20	10/03/2025	13:20	LB137426
	Silver	1.41			-10	10	10/03/2025	13:20	LB137426
	Sodium	127			0	0	10/03/2025	13:20	LB137426
	Thallium	11.3			-40	40	10/03/2025	13:20	LB137426
	Vanadium	4.84			-40	40	10/03/2025	13:20	LB137426
	Zinc	1.66			-40	40	10/03/2025	13:20	LB137426
ICSAB01	Aluminum	247000	247000	100	209000	285000	10/03/2025	13:25	LB137426
	Antimony	609	618	98	525	711	10/03/2025	13:25	LB137426
	Arsenic	110	104	106	88.4	120	10/03/2025	13:25	LB137426
	Barium	496	537	92	437	637	10/03/2025	13:25	LB137426
	Beryllium	516	495	104	420	570	10/03/2025	13:25	LB137426
	Cadmium	980	972	101	826	1120	10/03/2025	13:25	LB137426
	Calcium	235000	235000	100	199000	271000	10/03/2025	13:25	LB137426
	Chromium	553	542	102	460	624	10/03/2025	13:25	LB137426
	Cobalt	491	476	103	404	548	10/03/2025	13:25	LB137426
	Copper	484	511	95	434	588	10/03/2025	13:25	LB137426
	Iron	97500	99300	98	84400	114500	10/03/2025	13:25	LB137426
	Lead	49.1	49.0	100	37	61	10/03/2025	13:25	LB137426
	Magnesium	248000	248000	100	210000	286000	10/03/2025	13:25	LB137426
	Manganese	492	507	97	430	584	10/03/2025	13:25	LB137426
	Nickel	981	954	103	810	1100	10/03/2025	13:25	LB137426
	Potassium	15.2			0	0	10/03/2025	13:25	LB137426
	Selenium	56.4	46.0	123	26	66	10/03/2025	13:25	LB137426
	Silver	208	201	104	170	232	10/03/2025	13:25	LB137426
	Sodium	92.2			0	0	10/03/2025	13:25	LB137426
	Thallium	105	108	97	68	148	10/03/2025	13:25	LB137426

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>CDM Smith</u>	SDG No.: <u>Q3266</u>
Contract: <u>CAMP02</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	485	491	99	417	565	10/03/2025	13:25	LB137426
	Zinc	1030	952	108	809	1095	10/03/2025	13:25	LB137426
ICSA01	Aluminum	243000	255000	95	216000	294000	10/07/2025	12:14	LB137447
	Antimony	-5.05			-50	50	10/07/2025	12:14	LB137447
	Arsenic	3.53			-20	20	10/07/2025	12:14	LB137447
	Barium	-8.30	6.0	138	-94	106	10/07/2025	12:14	LB137447
	Beryllium	1.74			-6	6	10/07/2025	12:14	LB137447
	Cadmium	2.40	1.0	240	-5	7	10/07/2025	12:14	LB137447
	Calcium	234000	245000	96	208000	282000	10/07/2025	12:14	LB137447
	Chromium	47.2	52.0	91	42	62	10/07/2025	12:14	LB137447
	Cobalt	1.06			-30	30	10/07/2025	12:14	LB137447
	Copper	13.8	2.0	690	-18	22	10/07/2025	12:14	LB137447
	Iron	94400	101000	94	85600	116500	10/07/2025	12:14	LB137447
	Lead	-1.11			-12	12	10/07/2025	12:14	LB137447
	Magnesium	246000	255000	96	216000	294000	10/07/2025	12:14	LB137447
	Manganese	6.71	7.0	96	-13	27	10/07/2025	12:14	LB137447
	Nickel	2.60	2.0	130	-38	42	10/07/2025	12:14	LB137447
	Potassium	-102			0	0	10/07/2025	12:14	LB137447
	Selenium	8.06			-20	20	10/07/2025	12:14	LB137447
	Silver	-0.34			-10	10	10/07/2025	12:14	LB137447
	Sodium	-109			0	0	10/07/2025	12:14	LB137447
	Thallium	33.6			-40	40	10/07/2025	12:14	LB137447
	Vanadium	-0.97			-40	40	10/07/2025	12:14	LB137447
	Zinc	-0.92			-40	40	10/07/2025	12:14	LB137447
ICSAB01	Aluminum	247000	247000	100	209000	285000	10/07/2025	12:37	LB137447
	Antimony	604	618	98	525	711	10/07/2025	12:37	LB137447
	Arsenic	107	104	103	88.4	120	10/07/2025	12:37	LB137447
	Barium	477	537	89	437	637	10/07/2025	12:37	LB137447
	Beryllium	496	495	100	420	570	10/07/2025	12:37	LB137447
	Cadmium	990	972	102	826	1120	10/07/2025	12:37	LB137447
	Calcium	231000	235000	98	199000	271000	10/07/2025	12:37	LB137447
	Chromium	540	542	100	460	624	10/07/2025	12:37	LB137447
	Cobalt	493	476	104	404	548	10/07/2025	12:37	LB137447
	Copper	489	511	96	434	588	10/07/2025	12:37	LB137447
	Iron	97100	99300	98	84400	114500	10/07/2025	12:37	LB137447
	Lead	50.4	49.0	103	37	61	10/07/2025	12:37	LB137447
	Magnesium	246000	248000	99	210000	286000	10/07/2025	12:37	LB137447
	Manganese	477	507	94	430	584	10/07/2025	12:37	LB137447
	Nickel	982	954	103	810	1100	10/07/2025	12:37	LB137447
	Potassium	-317			0	0	10/07/2025	12:37	LB137447
	Selenium	60.1	46.0	131	26	66	10/07/2025	12:37	LB137447
	Silver	205	201	102	170	232	10/07/2025	12:37	LB137447

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: CDM Smith

SDG No.: Q3266

Contract: CAMP02

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Sodium	-119			0	0	10/07/2025	12:37	LB137447
	Thallium	140	108	130	68	148	10/07/2025	12:37	LB137447
	Vanadium	475	491	97	417	565	10/07/2025	12:37	LB137447
	Zinc	1020	952	107	809	1095	10/07/2025	12:37	LB137447