

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith
Contract: CAMP02

SDG No.: Q3266
Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169960BS							
Aluminum	mg/Kg	100	96.8		97	80 - 120	P
Antimony	mg/Kg	40.0	38.9		97	80 - 120	P
Arsenic	mg/Kg	40.0	38.7		97	80 - 120	P
Barium	mg/Kg	10.0	9.41		94	80 - 120	P
Beryllium	mg/Kg	10.0	10.3		103	80 - 120	P
Cadmium	mg/Kg	10.0	9.53		95	80 - 120	P
Calcium	mg/Kg	50.0	47.9	J	96	80 - 120	P
Chromium	mg/Kg	20.0	20.2		101	80 - 120	P
Cobalt	mg/Kg	10.0	9.54		95	80 - 120	P
Copper	mg/Kg	15.0	15.0		100	80 - 120	P
Iron	mg/Kg	150	146		97	80 - 120	P
Lead	mg/Kg	50.0	46.9		94	80 - 120	P
Magnesium	mg/Kg	100	98.8	J	99	80 - 120	P
Manganese	mg/Kg	10.0	9.71		97	80 - 120	P
Nickel	mg/Kg	25.0	24.4		98	80 - 120	P
Potassium	mg/Kg	500	437		87	80 - 120	P
Selenium	mg/Kg	100	98.9		99	80 - 120	P
Silver	mg/Kg	3.8	3.66		96	80 - 120	P
Sodium	mg/Kg	150	137		91	80 - 120	P
Thallium	mg/Kg	100	114		114	80 - 120	P
Vanadium	mg/Kg	15.0	14.5		97	80 - 120	P
Zinc	mg/Kg	10.0	9.84		98	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith

SDG No.: Q3266

Contract: CAMP02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169964BS Mercury	mg/Kg	0.25	0.27		106	80 - 120	CV