

Continuing Calibration Report for SequenceID : FD100225AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049798.D

Aromatic C10-C12	4619689.000	40.000	4.389	6.131	115492.225	122789.439	5.943
Aromatic C12-C16	8072201.000	60.000	6.132	8.759	134536.683	143358.701	6.154
Aromatic C16-C21	10659760.000	80.000	8.760	13.049	133247.000	142171.561	6.277
Aromatic C21-C36	20256445.000	180.000	13.050	18.468	112535.806	123821.788	9.115
Aromatic EPH	43608095.000	360.000	4.389	18.468	121133.597	131040.962	7.561