

Continuing Calibration Report for SequenceID : FD100325AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049814.D

Aromatic C10-C12	5171373.000	40.000	4.432	6.134	129284.325	122789.439	-5.289
Aromatic C12-C16	9033436.000	60.000	6.135	8.763	150557.267	143358.701	-5.021
Aromatic C16-C21	11767249.000	80.000	8.764	13.053	147090.613	142171.561	-3.460
Aromatic C21-C36	23491530.000	180.000	13.054	18.473	130508.500	123821.788	-5.400
Aromatic EPH	49463588.000	360.000	4.432	18.473	137398.856	131040.962	-4.852