

Continuing Calibration Report for SequenceID : FD100325AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049817.D

Aromatic C10-C12	4926316.000	40.000	4.432	6.134	123157.900	122789.439	-0.300
Aromatic C12-C16	8615482.000	60.000	6.135	8.763	143591.367	143358.701	-0.162
Aromatic C16-C21	11354359.000	80.000	8.764	13.053	141929.488	142171.561	0.170
Aromatic C21-C36	22775904.000	180.000	13.054	18.473	126532.800	123821.788	-2.189
Aromatic EPH	47672061.000	360.000	4.432	18.473	132422.392	131040.962	-1.054