

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name:	<u>Alliance</u>	Client:	<u>CDM Smith</u>
Lab Code:	<u>ACE</u>	SDG No:	<u>Q3266</u>
Sample No :	<u>PB169950BS</u>	Datafile:	<u>FE056148.D</u>
		Client ID :	<u>PB169950BS</u>

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS (%)
Aliphatic C9-C12	10	5.10	51		(40-140)
Aliphatic C12-C16	6.7	4.79	72		(40-140)
Aliphatic C16-C21	10	8.05	81		(40-140)
Aliphatic C21-C28	13.3	10.6	80		(40-140)
Aliphatic C28-C40	20.0	17.9	90		(40-140)