

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name:	<u>Alliance</u>	Client:	<u>CDM Smith</u>
Lab Code:	<u>ACE</u>	SDG No:	<u>Q3266</u>
Sample No :	<u>PB169950BSD</u>	Datafile:	<u>FE056149.D</u>
		Client ID :	<u>PB169950BSD</u>

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS (%)	QC Limit Of RPD
Aliphatic C9-C12	10	4.76	48		6.8 (40-140)	25
Aliphatic C12-C16	6.7	4.57	69		4.7 (40-140)	25
Aliphatic C16-C21	10	7.92	79		1.7 (40-140)	25
Aliphatic C21-C28	13.3	10.6	80		0.0257 (40-140)	25
Aliphatic C28-C40	20.0	17.7	89		1.1 (40-140)	25