

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3266
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/06/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031955.D **Time Analyzed:** 13:01

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.581	9.483	9.683	760.830	712.500	6.8
2,4,5-TP(Silvex)	9.283	9.185	9.385	762.520	712.500	7.0
2,4-D	8.414	8.314	8.514	743.080	705.000	5.4
2,4-DB	10.158	10.058	10.258	641.100	712.500	-10.0
2,4-DCAA	7.281	7.185	7.385	737.370	750.000	-1.7
DICAMBA	7.465	7.369	7.569	749.580	705.000	6.3
DICHLORPROP	8.174	8.077	8.277	724.470	705.000	2.8
Dinoseb	11.350	11.254	11.454	833.610	705.000	18.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3266
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/06/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031955.D **Time Analyzed:** 13:01

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.316	10.221	10.421	694.530	712.500	-2.5
2,4,5-TP(Silvex)	9.886	9.793	9.993	736.230	712.500	3.3
2,4-D	8.996	8.900	9.100	683.870	705.000	-3.0
2,4-DB	10.885	10.790	10.990	618.050	712.500	-13.3
2,4-DCAA	7.737	7.643	7.843	716.040	750.000	-4.5
DICAMBA	7.932	7.839	8.039	728.270	705.000	3.3
DICHLORPROP	8.652	8.559	8.759	736.180	705.000	4.4
Dinoseb	11.258	11.166	11.366	839.330	705.000	19.1