

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3266
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/06/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031968.D **Time Analyzed:** 21:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.576	9.483	9.683	794.500	712.500	11.5
2,4,5-TP(Silvex)	9.280	9.185	9.385	753.880	712.500	5.8
2,4-D	8.409	8.314	8.514	813.770	705.000	15.4
2,4-DB	10.152	10.058	10.258	792.790	712.500	11.3
2,4-DCAA	7.280	7.185	7.385	783.380	750.000	4.5
DICAMBA	7.464	7.369	7.569	742.370	705.000	5.3
DICHLORPROP	8.171	8.077	8.277	727.460	705.000	3.2
Dinoseb	11.345	11.254	11.454	787.330	705.000	11.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3266
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/06/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031968.D **Time Analyzed:** 21:26

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.318	10.221	10.421	727.990	712.500	2.2
2,4,5-TP(Silvex)	9.889	9.793	9.993	741.760	712.500	4.1
2,4-D	8.998	8.900	9.100	714.010	705.000	1.3
2,4-DB	10.886	10.790	10.990	632.870	712.500	-11.2
2,4-DCAA	7.741	7.643	7.843	754.270	750.000	0.6
DICAMBA	7.936	7.839	8.039	732.070	705.000	3.8
DICHLORPROP	8.655	8.559	8.759	744.960	705.000	5.7
Dinoseb	11.260	11.166	11.366	798.080	705.000	13.2