

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS090925

Review By	Abdul	Review On	9/10/2025 8:43:49 AM
Supervise By	mohammad	Supervise On	9/11/2025 1:29:37 AM
SubDirectory	PS090925	HP Acquire Method	HP Processing Method ps090925 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24556,PP24557,PP24558,PP24559,PP24560,PP24921
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24923
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS031666.D	09 Sep 2025 10:44		AR\AJ	Ok
2	I.BLK	I.BLK	PS031667.D	09 Sep 2025 11:08		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS031668.D	09 Sep 2025 11:32		AR\AJ	Ok,M
4	HSTDICC500	HSTDICC500	PS031669.D	09 Sep 2025 11:56		AR\AJ	Ok
5	HSTDICC750	HSTDICC750	PS031670.D	09 Sep 2025 12:20		AR\AJ	Ok
6	HSTDICC1000	HSTDICC1000	PS031671.D	09 Sep 2025 12:44		AR\AJ	Ok,M
7	HSTDICC1500	HSTDICC1500	PS031672.D	09 Sep 2025 13:08		AR\AJ	Ok,M
8	HSTDICV750	ICVPS090925	PS031673.D	09 Sep 2025 13:32		AR\AJ	Ok
9	I.BLK	I.BLK	PS031674.D	09 Sep 2025 13:57		AR\AJ	Ok
10	HSTDCCC750	HSTDCCC750	PS031675.D	09 Sep 2025 14:21		AR\AJ	Ok
11	PB169584BL	PB169584BL	PS031676.D	09 Sep 2025 14:45		AR\AJ	Ok
12	Q3015-17	WP0925-PT-HERB-WP	PS031677.D	09 Sep 2025 15:09		AR\AJ	Ok,M
13	PB169584BS	PB169584BS	PS031678.D	09 Sep 2025 15:33		AR\AJ	Ok
14	PB169584BSD	PB169584BSD	PS031679.D	09 Sep 2025 15:57		AR\AJ	Ok
15	PB169590BL	PB169590BL	PS031680.D	09 Sep 2025 16:22		AR\AJ	Ok
16	PB169590BS	PB169590BS	PS031681.D	09 Sep 2025 16:46		AR\AJ	Ok
17	I.BLK	I.BLK	PS031682.D	09 Sep 2025 17:10		AR\AJ	Ok,M
18	HSTDCCC750	HSTDCCC750	PS031683.D	09 Sep 2025 17:34		AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS090925

Review By	Abdul	Review On	9/10/2025 8:43:49 AM
Supervise By	mohammad	Supervise On	9/11/2025 1:29:37 AM
SubDirectory	PS090925	HP Acquire Method	HP Processing Method ps090925 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24556,PP24557,PP24558,PP24559,PP24560,PP24921		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24923		

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS100625

Review By	Abdul	Review On	10/7/2025 12:25:51 PM
Supervise By		Supervise On	
SubDirectory	PS100625	HP Acquire Method	HP Processing Method ps090925 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24556,PP24557,PP24558,PP24559,PP24560,PP24921		
CCC	PP24559		
Internal Standard/PEM			
ICV/I.BLK	PP24923		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS031953.D	06 Oct 2025 09:48		AR\AJ	Ok
2	I.BLK	I.BLK	PS031954.D	06 Oct 2025 10:15		AR\AJ	Ok,NS
3	HSTDCCC750	HSTDCCC750	PS031955.D	06 Oct 2025 13:01		AR\AJ	Ok,NS
4	PB169991BL	PB169991BL	PS031956.D	06 Oct 2025 16:36		AR\AJ	Ok
5	PB169991BS	PB169991BS	PS031957.D	06 Oct 2025 17:00		AR\AJ	Ok,NS
6	Q3266-01	ENV-3-6FT	PS031958.D	06 Oct 2025 17:24		AR\AJ	Ok,NS
7	Q3266-01MS	ENV-3-6FTMS	PS031959.D	06 Oct 2025 17:48	some compound recovery fail	AR\AJ	Ok,NS
8	Q3266-01MSD	ENV-3-6FTMSD	PS031960.D	06 Oct 2025 18:13	some compound recovery fail & RPD fail for DINOSEB	AR\AJ	Ok,NS
9	Q3266-02	ENV-4-6FT	PS031961.D	06 Oct 2025 18:37		AR\AJ	Ok
10	Q3269-01	EG1-TP1	PS031962.D	06 Oct 2025 19:01		AR\AJ	Ok,NS
11	Q3269-04	EG1-TP2	PS031963.D	06 Oct 2025 19:25		AR\AJ	Ok,NS
12	Q3289-01	VNJ-240	PS031964.D	06 Oct 2025 19:49		AR\AJ	Ok,NS
13	Q3289-03	VNJ-261	PS031965.D	06 Oct 2025 20:13		AR\AJ	Ok
14	Q3289-05	72-11966	PS031966.D	06 Oct 2025 20:37		AR\AJ	Ok,NS
15	I.BLK	I.BLK	PS031967.D	06 Oct 2025 21:02		AR\AJ	Ok
16	HSTDCCC750	HSTDCCC750	PS031968.D	06 Oct 2025 21:26		AR\AJ	Ok,NS

M : Manual Integration