

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3266</u>	Analytical Method:	<u>8151A</u>
Client:	<u>CDM Smith</u>	DataFile :	<u>PS031959.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q3266-01MS (Column 1)	Client Sample ID:	ENV-3-6FTMS									
	DICAMBA	207.1	0	93.9	ug/Kg	45				10	112	
	DICHLORPROP	207.1	0	104	ug/Kg	50				10	113	
	2,4-D	207.1	0	117	ug/Kg	56				10	144	
	2,4,5-TP(Silvex)	207.1	0	108	ug/Kg	52				10	114	
	2,4,5-T	207.1	0	108	ug/Kg	52				10	115	
	2,4-DB	207.1	0	93.1	ug/Kg	45				10	140	
	Dinoseb	207.1	0	0	ug/Kg	0	*			10	118	
Lab Sample ID:	Q3266-01MS (Column 2)	Client Sample ID:	ENV-3-6FTMS									
	DICAMBA	207.1	0	99.3	ug/Kg	48				10	112	
	DICHLORPROP	207.1	0	107	ug/Kg	52				10	113	
	2,4-D	207.1	0	122	ug/Kg	59				10	144	
	2,4,5-TP(Silvex)	207.1	0	115	ug/Kg	56				10	114	
	2,4,5-T	207.1	0	97.6	ug/Kg	47				10	115	
	2,4-DB	207.1	0	79.3	ug/Kg	38				10	140	
	Dinoseb	207.1	0	0	ug/Kg	0	*			10	118	