

**SW-846**

Datafile : PS031957.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits	
								Qual	Low	High
PB169991BS (Column 1)	DICAMBA	166.6	195	ug/Kg	117				72	129
	DICHLORPROP	166.6	192	ug/Kg	115				77	135
	2,4-D	166.6	207	ug/Kg	124				65	144
	2,4,5-TP(Silvex)	166.6	199	ug/Kg	119				74	146
	2,4,5-T	166.6	195	ug/Kg	117				77	134
	2,4-DB	166.6	161	ug/Kg	97				72	122
	Dinoseb	166.6	216	ug/Kg	130				74	132
PB169991BS (Column 2)	DICAMBA	166.6	188	ug/Kg	113				72	129
	DICHLORPROP	166.6	193	ug/Kg	116				77	135
	2,4-D	166.6	171	ug/Kg	103				65	144
	2,4,5-TP(Silvex)	166.6	192	ug/Kg	115				74	146
	2,4,5-T	166.6	175	ug/Kg	105				77	134
	2,4-DB	166.6	154	ug/Kg	92				72	122
	Dinoseb	166.6	214	ug/Kg	128				74	132