

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q3266

Continuing Calib Date: 10/02/2025

Initial Calibration Date(s): 09/23/2025

09/23/2025

Continuing Calib Time: 14:07

Initial Calibration Time(s): 08:45

17:14

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1242-1 (1)   | 5.80       | 5.81      | 5.71      | 5.91  | 0.01       |
| Aroclor-1242-2 (2)   | 5.82       | 5.83      | 5.73      | 5.93  | 0.01       |
| Aroclor-1242-3 (3)   | 5.89       | 5.90      | 5.80      | 6.00  | 0.01       |
| Aroclor-1242-4 (4)   | 5.98       | 5.99      | 5.89      | 6.09  | 0.01       |
| Aroclor-1242-5 (5)   | 6.71       | 6.72      | 6.62      | 6.82  | 0.01       |
| Tetrachloro-m-xylene | 4.65       | 4.66      | 4.56      | 4.76  | 0.01       |
| Decachlorobiphenyl   | 10.42      | 10.44     | 10.34     | 10.54 | 0.02       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q3266

Continuing Calib Date: 10/02/2025

Initial Calibration Date(s): 09/23/2025

09/23/2025

Continuing Calib Time: 14:07

Initial Calibration Time(s): 08:45

17:14

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1242-1 (1)   | 4.89       | 4.90      | 4.80      | 5.00 | 0.01       |
| Aroclor-1242-2 (2)   | 4.95       | 4.95      | 4.85      | 5.05 | 0.00       |
| Aroclor-1242-3 (3)   | 5.07       | 5.07      | 4.97      | 5.17 | 0.00       |
| Aroclor-1242-4 (4)   | 5.15       | 5.16      | 5.06      | 5.26 | 0.01       |
| Aroclor-1242-5 (5)   | 5.67       | 5.68      | 5.58      | 5.78 | 0.01       |
| Tetrachloro-m-xylene | 3.79       | 3.80      | 3.70      | 3.90 | 0.01       |
| Decachlorobiphenyl   | 8.79       | 8.81      | 8.71      | 8.91 | 0.02       |