

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q3266

Continuing Calib Date: 10/02/2025

Initial Calibration Date(s): 09/23/2025

09/23/2025

Continuing Calib Time: 14:39

Initial Calibration Time(s): 08:45

17:14

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.65	6.66	6.56	6.76	0.01
Aroclor-1254-2	(2)	6.87	6.88	6.78	6.98	0.02
Aroclor-1254-3	(3)	7.23	7.24	7.14	7.34	0.01
Aroclor-1254-4	(4)	7.51	7.52	7.42	7.62	0.01
Aroclor-1254-5	(5)	7.92	7.94	7.84	8.04	0.02
Tetrachloro-m-xylene		4.65	4.66	4.56	4.76	0.01
Decachlorobiphenyl		10.42	10.44	10.34	10.54	0.02

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q3266

Continuing Calib Date: 10/02/2025

Initial Calibration Date(s): 09/23/2025

09/23/2025

Continuing Calib Time: 14:39

Initial Calibration Time(s): 08:45

17:14

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.67	5.68	5.58	5.78	0.01
Aroclor-1254-2 (2)	5.82	5.83	5.73	5.93	0.01
Aroclor-1254-3 (3)	6.22	6.23	6.13	6.33	0.01
Aroclor-1254-4 (4)	6.44	6.46	6.36	6.56	0.02
Aroclor-1254-5 (5)	6.86	6.87	6.77	6.97	0.01
Tetrachloro-m-xylene	3.79	3.80	3.70	3.90	0.01
Decachlorobiphenyl	8.79	8.81	8.71	8.91	0.02