

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3266
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114211.D **Time Analyzed:** 14:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.736	4.643	4.843	564.240	500.000	12.8
Aroclor-1016-2	4.754	4.661	4.861	576.990	500.000	15.4
Aroclor-1016-3	4.812	4.718	4.918	551.890	500.000	10.4
Aroclor-1016-4	4.931	4.837	5.037	543.380	500.000	8.7
Aroclor-1016-5	5.187	5.093	5.293	595.440	500.000	19.1
Aroclor-1260-1	6.220	6.129	6.329	584.940	500.000	17.0
Aroclor-1260-2	6.411	6.319	6.519	581.610	500.000	16.3
Aroclor-1260-3	6.775	6.683	6.883	559.110	500.000	11.8
Aroclor-1260-4	7.032	6.941	7.141	606.930	500.000	21.4
Aroclor-1260-5	7.276	7.185	7.385	607.240	500.000	21.4
Decachlorobiphenyl	8.658	8.570	8.770	45.820	50.000	-8.4
Tetrachloro-m-xylene	3.656	3.560	3.760	47.190	50.000	-5.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3266
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114211.D **Time Analyzed:** 14:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.716	4.623	4.823	528.180	500.000	5.6
Aroclor-1016-2	4.734	4.641	4.841	525.050	500.000	5.0
Aroclor-1016-3	4.909	4.816	5.016	517.130	500.000	3.4
Aroclor-1016-4	4.951	4.858	5.058	527.500	500.000	5.5
Aroclor-1016-5	5.162	5.070	5.270	542.240	500.000	8.4
Aroclor-1260-1	6.187	6.096	6.296	496.030	500.000	-0.8
Aroclor-1260-2	6.376	6.284	6.484	465.980	500.000	-6.8
Aroclor-1260-3	6.526	6.435	6.635	456.650	500.000	-8.7
Aroclor-1260-4	6.994	6.903	7.103	461.550	500.000	-7.7
Aroclor-1260-5	7.237	7.146	7.346	454.240	500.000	-9.2
Decachlorobiphenyl	8.603	8.514	8.714	44.690	50.000	-10.6
Tetrachloro-m-xylene	3.647	3.551	3.751	43.730	50.000	-12.5