

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3266
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114223.D **Time Analyzed:** 19:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.736	4.643	4.843	473.500	500.000	-5.3
Aroclor-1016-2	4.754	4.661	4.861	478.300	500.000	-4.3
Aroclor-1016-3	4.811	4.718	4.918	464.270	500.000	-7.1
Aroclor-1016-4	4.930	4.837	5.037	442.510	500.000	-11.5
Aroclor-1016-5	5.186	5.093	5.293	489.330	500.000	-2.1
Aroclor-1260-1	6.221	6.129	6.329	510.150	500.000	2.0
Aroclor-1260-2	6.410	6.319	6.519	493.910	500.000	-1.2
Aroclor-1260-3	6.775	6.683	6.883	494.920	500.000	-1.0
Aroclor-1260-4	7.033	6.941	7.141	545.900	500.000	9.2
Aroclor-1260-5	7.277	7.185	7.385	555.580	500.000	11.1
Decachlorobiphenyl	8.657	8.570	8.770	51.520	50.000	3.0
Tetrachloro-m-xylene	3.656	3.560	3.760	50.040	50.000	0.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3266
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114223.D **Time Analyzed:** 19:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.715	4.623	4.823	419.850	500.000	-16.0
Aroclor-1016-2	4.733	4.641	4.841	428.210	500.000	-14.4
Aroclor-1016-3	4.907	4.816	5.016	437.230	500.000	-12.6
Aroclor-1016-4	4.950	4.858	5.058	444.220	500.000	-11.2
Aroclor-1016-5	5.161	5.070	5.270	457.440	500.000	-8.5
Aroclor-1260-1	6.186	6.096	6.296	428.640	500.000	-14.3
Aroclor-1260-2	6.375	6.284	6.484	405.950	500.000	-18.8
Aroclor-1260-3	6.525	6.435	6.635	401.260	500.000	-19.7
Aroclor-1260-4	6.993	6.903	7.103	401.680	500.000	-19.7
Aroclor-1260-5	7.236	7.146	7.346	401.700	500.000	-19.7
Decachlorobiphenyl	8.603	8.514	8.714	49.850	50.000	-0.3
Tetrachloro-m-xylene	3.646	3.551	3.751	47.260	50.000	-5.5