

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3266
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075540.D **Time Analyzed:** 13:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.803	5.708	5.908	422.820	500.000	-15.4
Aroclor-1016-2	5.825	5.729	5.929	422.540	500.000	-15.5
Aroclor-1016-3	5.888	5.792	5.992	414.640	500.000	-17.1
Aroclor-1016-4	5.985	5.889	6.089	451.120	500.000	-9.8
Aroclor-1016-5	6.277	6.181	6.381	449.960	500.000	-10.0
Aroclor-1260-1	7.393	7.299	7.499	422.190	500.000	-15.6
Aroclor-1260-2	7.646	7.551	7.751	429.750	500.000	-14.1
Aroclor-1260-3	8.004	7.909	8.109	449.230	500.000	-10.2
Aroclor-1260-4	8.232	8.137	8.337	496.470	500.000	-0.7
Aroclor-1260-5	8.556	8.464	8.664	480.500	500.000	-3.9
Decachlorobiphenyl	10.422	10.334	10.534	47.390	50.000	-5.2
Tetrachloro-m-xylene	4.653	4.555	4.755	44.820	50.000	-10.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3266
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075540.D **Time Analyzed:** 13:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.890	4.793	4.993	502.600	500.000	0.5
Aroclor-1016-2	4.947	4.851	5.051	518.860	500.000	3.8
Aroclor-1016-3	5.067	4.971	5.171	472.820	500.000	-5.4
Aroclor-1016-4	5.107	5.011	5.211	511.470	500.000	2.3
Aroclor-1016-5	5.322	5.225	5.425	489.060	500.000	-2.2
Aroclor-1260-1	6.349	6.255	6.455	439.880	500.000	-12.0
Aroclor-1260-2	6.536	6.441	6.641	438.310	500.000	-12.3
Aroclor-1260-3	6.689	6.595	6.795	438.790	500.000	-12.2
Aroclor-1260-4	7.158	7.065	7.265	402.320	500.000	-19.5
Aroclor-1260-5	7.397	7.304	7.504	423.530	500.000	-15.3
Decachlorobiphenyl	8.793	8.703	8.903	56.260	50.000	12.5
Tetrachloro-m-xylene	3.795	3.696	3.896	50.540	50.000	1.1