

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3266
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075555.D **Time Analyzed:** 18:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.800	5.708	5.908	437.650	500.000	-12.5
Aroclor-1016-2	5.821	5.729	5.929	435.390	500.000	-12.9
Aroclor-1016-3	5.884	5.792	5.992	425.240	500.000	-15.0
Aroclor-1016-4	5.981	5.889	6.089	457.280	500.000	-8.5
Aroclor-1016-5	6.274	6.181	6.381	465.980	500.000	-6.8
Aroclor-1260-1	7.392	7.299	7.499	441.590	500.000	-11.7
Aroclor-1260-2	7.645	7.551	7.751	441.100	500.000	-11.8
Aroclor-1260-3	8.003	7.909	8.109	470.750	500.000	-5.9
Aroclor-1260-4	8.231	8.137	8.337	495.580	500.000	-0.9
Aroclor-1260-5	8.556	8.464	8.664	493.160	500.000	-1.4
Decachlorobiphenyl	10.423	10.334	10.534	47.480	50.000	-5.0
Tetrachloro-m-xylene	4.648	4.555	4.755	45.490	50.000	-9.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3266
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075555.D **Time Analyzed:** 18:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.889	4.793	4.993	520.140	500.000	4.0
Aroclor-1016-2	4.947	4.851	5.051	528.500	500.000	5.7
Aroclor-1016-3	5.067	4.971	5.171	487.550	500.000	-2.5
Aroclor-1016-4	5.107	5.011	5.211	514.020	500.000	2.8
Aroclor-1016-5	5.322	5.225	5.425	500.490	500.000	0.1
Aroclor-1260-1	6.349	6.255	6.455	447.630	500.000	-10.5
Aroclor-1260-2	6.536	6.441	6.641	444.150	500.000	-11.2
Aroclor-1260-3	6.689	6.595	6.795	436.760	500.000	-12.6
Aroclor-1260-4	7.158	7.065	7.265	415.570	500.000	-16.9
Aroclor-1260-5	7.398	7.304	7.504	430.590	500.000	-13.9
Decachlorobiphenyl	8.795	8.703	8.903	54.930	50.000	9.9
Tetrachloro-m-xylene	3.795	3.696	3.896	51.180	50.000	2.4