

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO091525

Review By	yogesh	Review On	9/15/2025 12:11:18 PM
Supervise By	mohammad	Supervise On	9/17/2025 2:07:54 AM
SubDirectory	PO091525	HP Acquire Method	HP Processing Method PO091525

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO113718.D	15 Sep 2025 08:58		YP/AJ	Ok
2	I.BLK	I.BLK	PO113719.D	15 Sep 2025 09:16		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO113720.D	15 Sep 2025 09:34		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO113721.D	15 Sep 2025 09:53		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO113722.D	15 Sep 2025 10:11		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO113723.D	15 Sep 2025 10:30		YP/AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PO113724.D	15 Sep 2025 10:48		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO113725.D	15 Sep 2025 11:06		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO113726.D	15 Sep 2025 11:25		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO113727.D	15 Sep 2025 11:43		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO113728.D	15 Sep 2025 12:02		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO113729.D	15 Sep 2025 12:20		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO113730.D	15 Sep 2025 12:38		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO113731.D	15 Sep 2025 13:34		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO113732.D	15 Sep 2025 13:54		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO113733.D	15 Sep 2025 14:13		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO113734.D	15 Sep 2025 14:31		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO113735.D	15 Sep 2025 14:49		YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO091525

Review By	yogesh	Review On	9/15/2025 12:11:18 PM
Supervise By	mohammad	Supervise On	9/17/2025 2:07:54 AM
SubDirectory	PO091525	HP Acquire Method	HP Processing Method PO091525
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PO113736.D	15 Sep 2025 15:26		YP/AJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PO113737.D	15 Sep 2025 15:45		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO113738.D	15 Sep 2025 16:02		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO113739.D	15 Sep 2025 16:20		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO113740.D	15 Sep 2025 16:38		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO113741.D	15 Sep 2025 17:14		YP/AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PO113742.D	15 Sep 2025 17:33		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO113743.D	15 Sep 2025 17:51		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO113744.D	15 Sep 2025 18:10		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO113745.D	15 Sep 2025 18:27		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO113746.D	15 Sep 2025 18:45		YP/AJ	Ok
30	AR1268ICC050	AR1268ICC050	PO113747.D	15 Sep 2025 19:21		YP/AJ	Ok,M
31	PO091525ICV500	ICVPO091525	PO113748.D	15 Sep 2025 19:40		YP/AJ	Ok,M
32	AR1242ICV500	ICVPO091525AR1242	PO113749.D	15 Sep 2025 20:16		YP/AJ	Ok
33	AR1248ICV500	ICVPO091525AR1248	PO113750.D	15 Sep 2025 21:11		YP/AJ	Ok
34	AR1254ICV500	ICVPO091525AR1254	PO113751.D	15 Sep 2025 22:05		YP/AJ	Ok
35	AR1268ICV500	ICVPO091525AR1268	PO113752.D	15 Sep 2025 23:00		YP/AJ	Ok
36	DDT ANALOG	DDT ANALOG	PO113753.D	15 Sep 2025 23:37		YP/AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO100225

Review By	yogesh	Review On	10/2/2025 11:30:42 AM
Supervise By	mohammad	Supervise On	10/4/2025 6:31:26 AM
SubDirectory	PO100225	HP Acquire Method	HP Processing Method PO091925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO114200.D	02 Oct 2025 09:18		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO114201.D	02 Oct 2025 09:36		YP/AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PO114202.D	02 Oct 2025 10:06		YP/AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PO114203.D	02 Oct 2025 10:24		YP/AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PO114204.D	02 Oct 2025 10:43		YP/AJ	Ok,M
6	I.BLK	I.BLK	PO114205.D	02 Oct 2025 11:01		YP/AJ	Ok
7	Q3264-25	245F46-2-2	PO114206.D	02 Oct 2025 11:43		YP/AJ	Ok
8	Q3265-01	BC279304-LOZ-END-1	PO114207.D	02 Oct 2025 12:00		YP/AJ	Ok
9	Q3265-02	BC279304-LOZ-END-1	PO114208.D	02 Oct 2025 12:19		YP/AJ	Ok
10	PB169948BL	PB169948BL	PO114209.D	02 Oct 2025 12:55		YP/AJ	Ok
11	PB169948BS	PB169948BS	PO114210.D	02 Oct 2025 13:13		YP/AJ	Ok,M
12	AR1660CCC500	AR1660CCC500	PO114211.D	02 Oct 2025 14:28		YP/AJ	Ok,M
13	AR1242CCC500	AR1242CCC500	PO114212.D	02 Oct 2025 14:47		YP/AJ	Ok,M
14	AR1248CCC500	AR1248CCC500	PO114213.D	02 Oct 2025 15:05		YP/AJ	Ok
15	AR1254CCC500	AR1254CCC500	PO114214.D	02 Oct 2025 15:23		YP/AJ	Ok
16	I.BLK	I.BLK	PO114215.D	02 Oct 2025 15:42		YP/AJ	Ok
17	PB169951BL	PB169951BL	PO114216.D	02 Oct 2025 16:00		YP/AJ	Ok
18	PB169951BS	PB169951BS	PO114217.D	02 Oct 2025 16:19		YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO100225

Review By	yogesh	Review On	10/2/2025 11:30:42 AM
Supervise By	mohammad	Supervise On	10/4/2025 6:31:26 AM
SubDirectory	PO100225	HP Acquire Method	HP Processing Method PO091925
STD. NAME	STD REF.#		
Tune/Reschk	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q3262-08	279887	PO114218.D	02 Oct 2025 16:37		YP/AJ	Ok,M
20	Q3262-08MS	279887MS	PO114219.D	02 Oct 2025 16:55		YP/AJ	Ok,M
21	Q3262-08MSD	279887MSD	PO114220.D	02 Oct 2025 17:14		YP/AJ	Ok,M
22	Q3249-08	S-8(4.5-5)	PO114221.D	02 Oct 2025 17:32		YP/AJ	Ok
23	Q3249-16	S-14(4-4.5)	PO114222.D	02 Oct 2025 17:51		YP/AJ	Ok,M
24	AR1660CCC500	AR1660CCC500	PO114223.D	02 Oct 2025 19:24		YP/AJ	Ok,M
25	AR1242CCC500	AR1242CCC500	PO114224.D	02 Oct 2025 20:01		YP/AJ	Ok,M
26	AR1248CCC500	AR1248CCC500	PO114225.D	02 Oct 2025 20:20		YP/AJ	Ok,M
27	AR1254CCC500	AR1254CCC500	PO114226.D	02 Oct 2025 20:38		YP/AJ	Ok,M
28	I.BLK	I.BLK	PO114227.D	02 Oct 2025 20:56		YP/AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP092325

Review By	yogesh	Review On	9/23/2025 12:35:13 PM
Supervise By	mohammad	Supervise On	9/25/2025 2:25:37 AM
SubDirectory	PP092325	HP Acquire Method	HP Processing Method PP092325

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP075245.D	23 Sep 2025 08:13		YP\AJ	Ok
2	I.BLK	I.BLK	PP075246.D	23 Sep 2025 08:29		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP075247.D	23 Sep 2025 08:45		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP075248.D	23 Sep 2025 09:01		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP075249.D	23 Sep 2025 09:18		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP075250.D	23 Sep 2025 09:34		YP\AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PP075251.D	23 Sep 2025 10:11		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP075252.D	23 Sep 2025 10:27		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP075253.D	23 Sep 2025 10:43		YP\AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP075254.D	23 Sep 2025 11:00		YP\AJ	Ok,M
11	AR1242ICC750	AR1242ICC750	PP075255.D	23 Sep 2025 11:16		YP\AJ	Ok,M
12	AR1242ICC500	AR1242ICC500	PP075256.D	23 Sep 2025 11:32		YP\AJ	Ok,M
13	AR1242ICC250	AR1242ICC250	PP075257.D	23 Sep 2025 11:48		YP\AJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PP075258.D	23 Sep 2025 12:05		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP075259.D	23 Sep 2025 12:37		YP\AJ	Ok,M
16	AR1248ICC750	AR1248ICC750	PP075260.D	23 Sep 2025 12:53		YP\AJ	Ok,M
17	AR1248ICC500	AR1248ICC500	PP075261.D	23 Sep 2025 13:10		YP\AJ	Ok,M
18	AR1248ICC250	AR1248ICC250	PP075262.D	23 Sep 2025 13:26		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP092325

Review By	yogesh	Review On	9/23/2025 12:35:13 PM
Supervise By	mohammad	Supervise On	9/25/2025 2:25:37 AM
SubDirectory	PP092325	HP Acquire Method	HP Processing Method PP092325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PP075263.D	23 Sep 2025 13:42		YPIAJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP075264.D	23 Sep 2025 14:15		YPIAJ	Ok,M
21	AR1254ICC750	AR1254ICC750	PP075265.D	23 Sep 2025 14:31		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP075266.D	23 Sep 2025 14:47		YPIAJ	Ok,M
23	AR1254ICC250	AR1254ICC250	PP075267.D	23 Sep 2025 15:04		YPIAJ	Ok,M
24	AR1254ICC050	AR1254ICC050	PP075268.D	23 Sep 2025 15:20		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP075269.D	23 Sep 2025 15:52		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP075270.D	23 Sep 2025 16:09		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PP075271.D	23 Sep 2025 16:25		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PP075272.D	23 Sep 2025 16:41		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP075273.D	23 Sep 2025 16:58		YPIAJ	Ok
30	AR1268ICC050	AR1268ICC050	PP075274.D	23 Sep 2025 17:14		YPIAJ	Ok,M
31	PP092325ICV500	ICVPP092325	PP075275.D	23 Sep 2025 17:46		YPIAJ	Ok,M
32	AR1242ICV500	ICVPP092325AR1242	PP075276.D	23 Sep 2025 18:35		YPIAJ	Ok,M
33	AR1248ICV500	ICVPP092325AR1248	PP075277.D	23 Sep 2025 19:08		YPIAJ	Ok,M
34	AR1254ICV500	ICVPP092325AR1254	PP075278.D	23 Sep 2025 19:40		YPIAJ	Ok,M
35	AR1268ICV500	ICVPP092325AR1268	PP075279.D	23 Sep 2025 19:57		YPIAJ	Ok,M
36	DDT ANALOG	DDT ANALOG	PP075280.D	23 Sep 2025 20:29		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP100225

Review By	yogesh	Review On	10/2/2025 11:32:00 AM
Supervise By	mohammad	Supervise On	10/4/2025 6:31:54 AM
SubDirectory	PP100225	HP Acquire Method	HP Processing Method PP092325

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP075529.D	02 Oct 2025 09:14		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP075530.D	02 Oct 2025 09:30		YP\AJ	Ok
3	AR1242CCC500	AR1242CCC500	PP075531.D	02 Oct 2025 10:01		YP\AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP075532.D	02 Oct 2025 10:17		YP\AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP075533.D	02 Oct 2025 10:33		YP\AJ	Ok,M
6	I.BLK	I.BLK	PP075534.D	02 Oct 2025 10:49		YP\AJ	Ok
7	Q3264-21	245F55-1-2	PP075535.D	02 Oct 2025 11:41		YP\AJ	Ok,M
8	Q3264-22	245F46-1-1	PP075536.D	02 Oct 2025 11:57		YP\AJ	Ok,M
9	Q3264-23	245F46-1-2	PP075537.D	02 Oct 2025 12:13	Surroage fail	YP\AJ	ReRun
10	Q3264-24	245F46-2-1	PP075538.D	02 Oct 2025 12:29		YP\AJ	Ok,M
11	Q3264-23RE	245F46-1-2	PP075539.D	02 Oct 2025 12:46	Surroage fail	YP\AJ	Confirms
12	AR1660CCC500	AR1660CCC500	PP075540.D	02 Oct 2025 13:51		YP\AJ	Ok
13	AR1242CCC500	AR1242CCC500	PP075541.D	02 Oct 2025 14:07		YP\AJ	Ok,M
14	AR1248CCC500	AR1248CCC500	PP075542.D	02 Oct 2025 14:23		YP\AJ	Ok,M
15	AR1254CCC500	AR1254CCC500	PP075543.D	02 Oct 2025 14:39		YP\AJ	Ok,M
16	I.BLK	I.BLK	PP075544.D	02 Oct 2025 14:56		YP\AJ	Ok
17	Q3262-01	PAD-100125	PP075545.D	02 Oct 2025 15:12		YP\AJ	Ok
18	Q3262-02	MOO-25-0280	PP075546.D	02 Oct 2025 15:28		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP100225

Review By	yogesh	Review On	10/2/2025 11:32:00 AM
Supervise By	mohammad	Supervise On	10/4/2025 6:31:54 AM
SubDirectory	PP100225	HP Acquire Method	HP Processing Method PP092325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q3262-05	MOO-25-0281-82	PP075547.D	02 Oct 2025 15:44		YPIAJ	Ok,M
20	Q3266-01	ENV-3-6FT	PP075548.D	02 Oct 2025 16:01		YPIAJ	Ok
21	Q3266-02	ENV-4-6FT	PP075549.D	02 Oct 2025 16:17	AR1254 Hit	YPIAJ	Ok,M
22	Q3249-03	S-3(4-4.5)	PP075550.D	02 Oct 2025 16:33	AR1254 & AR1260 hits	YPIAJ	Ok
23	Q3249-19	S-18(4.5-5)	PP075551.D	02 Oct 2025 16:49	AR1254 & AR1260 hits	YPIAJ	Ok
24	Q3249-22	S-20(4-4.5)	PP075552.D	02 Oct 2025 17:06	AR1254 & AR1260 hits, Need 10X	YPIAJ	Dilution
25	Q3249-24	S-22(4.5-5)	PP075553.D	02 Oct 2025 17:22	AR1254 & AR1260 hits, Need 25X	YPIAJ	Dilution
26	Q3249-27	S-24(4.5-5)	PP075554.D	02 Oct 2025 17:38	AR1260 & AR1268 hits, Need to confirm hit, AR1268 ccal missing, Need 2X	YPIAJ	Not Ok
27	AR1660CCC500	AR1660CCC500	PP075555.D	02 Oct 2025 18:59		YPIAJ	Ok
28	AR1242CCC500	AR1242CCC500	PP075556.D	02 Oct 2025 19:48		YPIAJ	Ok,M
29	AR1248CCC500	AR1248CCC500	PP075557.D	02 Oct 2025 20:04		YPIAJ	Ok,M
30	AR1254CCC500	AR1254CCC500	PP075558.D	02 Oct 2025 20:20		YPIAJ	Ok,M
31	I.BLK	I.BLK	PP075559.D	02 Oct 2025 20:37		YPIAJ	Ok

M : Manual Integration