

Surrogate Summary

SDG No.: Q3266

Client: CDM Smith

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO113719.D	PIBLK-PO113719.D	Tetrachloro-m-xyl	1	20	17.0	85		60	140
		Decachlorobiphen	1	20	17.3	86		60	140
		Tetrachloro-m-xyl	2	20	17.0	85		60	140
		Decachlorobiphen	2	20	18.1	91		60	140
I.BLK-PO114215.D	PIBLK-PO114215.D	Tetrachloro-m-xyl	1	20	16.5	83		60	140
		Decachlorobiphen	1	20	17.2	86		60	140
		Tetrachloro-m-xyl	2	20	15.7	79		60	140
		Decachlorobiphen	2	20	16.7	83		60	140
PB169951BL	PB169951BL	Tetrachloro-m-xyl	1	20	17.9	90		32	144
		Decachlorobiphen	1	20	18.8	94		32	175
		Tetrachloro-m-xyl	2	20	17.1	85		32	144
		Decachlorobiphen	2	20	18.2	91		32	175
PB169951BS	PB169951BS	Tetrachloro-m-xyl	1	20	18.0	90		32	144
		Decachlorobiphen	1	20	19.5	98		32	175
		Tetrachloro-m-xyl	2	20	17.1	85		32	144
		Decachlorobiphen	2	20	19.1	96		32	175
Q3262-08MS	279887MS	Tetrachloro-m-xyl	1	20	9.38	47		32	144
		Decachlorobiphen	1	20	18.9	94		32	175
		Tetrachloro-m-xyl	2	20	9.97	50		32	144
		Decachlorobiphen	2	20	19.4	97		32	175
Q3262-08MSD	279887MSD	Tetrachloro-m-xyl	1	20	9.55	48		32	144
		Decachlorobiphen	1	20	19.2	96		32	175
		Tetrachloro-m-xyl	2	20	10.3	51		32	144
		Decachlorobiphen	2	20	19.6	98		32	175
I.BLK-PO114227.D	PIBLK-PO114227.D	Tetrachloro-m-xyl	1	20	16.7	84		60	140
		Decachlorobiphen	1	20	18.4	92		60	140
		Tetrachloro-m-xyl	2	20	16.0	80		60	140
		Decachlorobiphen	2	20	17.5	87		60	140
I.BLK-PP075246.D	PIBLK-PP075246.D	Tetrachloro-m-xyl	1	20	19.9	100		60	140
		Decachlorobiphen	1	20	19.4	97		60	140
		Tetrachloro-m-xyl	2	20	17.9	90		60	140
		Decachlorobiphen	2	20	20.0	100		60	140
I.BLK-PP075544.D	PIBLK-PP075544.D	Tetrachloro-m-xyl	1	20	16.7	84		60	140
		Decachlorobiphen	1	20	19.5	98		60	140
		Tetrachloro-m-xyl	2	20	15.1	75		60	140
		Decachlorobiphen	2	20	21.0	105		60	140
Q3266-01	ENV-3-6FT	Tetrachloro-m-xyl	1	20	15.6	78		32	144
		Decachlorobiphen	1	20	12.5	62		32	175
		Tetrachloro-m-xyl	2	20	14.2	71		32	144
		Decachlorobiphen	2	20	12.1	61		32	175
Q3266-02	ENV-4-6FT	Tetrachloro-m-xyl	1	20	17.0	85		32	144

Surrogate Summary

SDG No.: Q3266

Client: CDM Smith

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q3266-02	ENV-4-6FT	Decachlorobiphen	1	20	14.8	74		32	175
		Tetrachloro-m-xyl	2	20	16.6	83		32	144
		Decachlorobiphen	2	20	15.2	76		32	175
I.BLK-PP075559.D	PIBLK-PP075559.D	Tetrachloro-m-xyl	1	20	17.5	87		60	140
		Decachlorobiphen	1	20	19.3	96		60	140
		Tetrachloro-m-xyl	2	20	17.2	86		60	140
		Decachlorobiphen	2	20	20.0	100		60	140