

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3266
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097495.D **Time Analyzed:** 09:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.660	6.566	6.766	54.930	50.000	9.9
4,4'-DDE	6.151	6.056	6.256	50.790	50.000	1.6
4,4'-DDT	6.975	6.880	7.080	48.520	50.000	-3.0
Aldrin	5.229	5.132	5.332	47.320	50.000	-5.4
alpha-BHC	3.972	3.875	4.075	47.820	50.000	-4.4
alpha-Chlordane	5.981	5.886	6.086	48.300	50.000	-3.4
beta-BHC	4.487	4.390	4.590	47.260	50.000	-5.5
Decachlorobiphenyl	8.994	8.902	9.102	46.370	50.000	-7.3
delta-BHC	4.732	4.635	4.835	49.480	50.000	-1.0
Dieldrin	6.302	6.206	6.406	49.530	50.000	-0.9
Endosulfan I	6.028	5.934	6.134	48.640	50.000	-2.7
Endosulfan II	6.740	6.646	6.846	46.720	50.000	-6.6
Endosulfan sulfate	7.104	7.007	7.207	51.910	50.000	3.8
Endrin	6.528	6.432	6.632	50.200	50.000	0.4
Endrin aldehyde	6.869	6.774	6.974	54.440	50.000	8.9
Endrin ketone	7.582	7.487	7.687	54.700	50.000	9.4
gamma-BHC (Lindane)	4.300	4.202	4.402	47.270	50.000	-5.5
gamma-Chlordane	5.901	5.806	6.006	48.480	50.000	-3.0
Heptachlor	4.890	4.793	4.993	45.680	50.000	-8.6
Heptachlor epoxide	5.647	5.552	5.752	49.050	50.000	-1.9
Methoxychlor	7.448	7.352	7.552	46.010	50.000	-8.0
Tetrachloro-m-xylene	3.526	3.428	3.628	44.030	50.000	-11.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3266
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097495.D **Time Analyzed:** 09:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.846	5.750	5.950	48.040	50.000	-3.9
4,4'-DDE	5.294	5.197	5.397	45.540	50.000	-8.9
4,4'-DDT	6.098	6.002	6.202	40.190	50.000	-19.6
Aldrin	4.288	4.191	4.391	46.450	50.000	-7.1
alpha-BHC	3.328	3.229	3.429	47.080	50.000	-5.8
alpha-Chlordane	5.106	5.009	5.209	46.800	50.000	-6.4
beta-BHC	3.955	3.857	4.057	43.640	50.000	-12.7
Decachlorobiphenyl	7.979	7.884	8.084	41.750	50.000	-16.5
delta-BHC	4.187	4.090	4.290	45.550	50.000	-8.9
Dieldrin	5.423	5.328	5.528	47.410	50.000	-5.2
Endosulfan I	5.160	5.064	5.264	42.580	50.000	-14.8
Endosulfan II	5.991	5.895	6.095	46.500	50.000	-7.0
Endosulfan sulfate	6.392	6.296	6.496	46.530	50.000	-6.9
Endrin	5.700	5.603	5.803	53.130	50.000	6.3
Endrin aldehyde	6.169	6.073	6.273	45.750	50.000	-8.5
Endrin ketone	6.895	6.800	7.000	51.220	50.000	2.4
gamma-BHC (Lindane)	3.659	3.561	3.761	46.170	50.000	-7.7
gamma-Chlordane	5.041	4.945	5.145	46.270	50.000	-7.5
Heptachlor	4.007	3.909	4.109	44.450	50.000	-11.1
Heptachlor epoxide	4.790	4.693	4.893	46.620	50.000	-6.8
Methoxychlor	6.670	6.574	6.774	40.370	50.000	-19.3
Tetrachloro-m-xylene	2.823	2.724	2.924	41.980	50.000	-16.0