

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3266
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097506.D **Time Analyzed:** 16:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.660	6.566	6.766	56.870	50.000	13.7
4,4'-DDE	6.153	6.056	6.256	52.110	50.000	4.2
4,4'-DDT	6.976	6.880	7.080	51.160	50.000	2.3
Aldrin	5.229	5.132	5.332	48.390	50.000	-3.2
alpha-BHC	3.972	3.875	4.075	49.320	50.000	-1.4
alpha-Chlordane	5.981	5.886	6.086	49.760	50.000	-0.5
beta-BHC	4.487	4.390	4.590	48.450	50.000	-3.1
Decachlorobiphenyl	8.995	8.902	9.102	47.770	50.000	-4.5
delta-BHC	4.731	4.635	4.835	50.900	50.000	1.8
Dieldrin	6.302	6.206	6.406	51.120	50.000	2.2
Endosulfan I	6.030	5.934	6.134	51.800	50.000	3.6
Endosulfan II	6.740	6.646	6.846	48.610	50.000	-2.8
Endosulfan sulfate	7.104	7.007	7.207	53.040	50.000	6.1
Endrin	6.528	6.432	6.632	52.040	50.000	4.1
Endrin aldehyde	6.869	6.774	6.974	56.070	50.000	12.1
Endrin ketone	7.582	7.487	7.687	55.730	50.000	11.5
gamma-BHC (Lindane)	4.300	4.202	4.402	48.960	50.000	-2.1
gamma-Chlordane	5.901	5.806	6.006	50.180	50.000	0.4
Heptachlor	4.890	4.793	4.993	47.100	50.000	-5.8
Heptachlor epoxide	5.647	5.552	5.752	51.060	50.000	2.1
Methoxychlor	7.448	7.352	7.552	48.850	50.000	-2.3
Tetrachloro-m-xylene	3.525	3.428	3.628	45.090	50.000	-9.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3266
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097506.D **Time Analyzed:** 16:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.846	5.750	5.950	48.480	50.000	-3.0
4,4'-DDE	5.294	5.197	5.397	45.720	50.000	-8.6
4,4'-DDT	6.098	6.002	6.202	40.980	50.000	-18.0
Aldrin	4.289	4.191	4.391	46.840	50.000	-6.3
alpha-BHC	3.328	3.229	3.429	47.550	50.000	-4.9
alpha-Chlordane	5.106	5.009	5.209	47.090	50.000	-5.8
beta-BHC	3.955	3.857	4.057	44.210	50.000	-11.6
Decachlorobiphenyl	7.979	7.884	8.084	41.540	50.000	-16.9
delta-BHC	4.188	4.090	4.290	46.090	50.000	-7.8
Dieldrin	5.425	5.328	5.528	48.900	50.000	-2.2
Endosulfan I	5.161	5.064	5.264	42.820	50.000	-14.4
Endosulfan II	5.991	5.895	6.095	46.670	50.000	-6.7
Endosulfan sulfate	6.392	6.296	6.496	46.810	50.000	-6.4
Endrin	5.700	5.603	5.803	54.880	50.000	9.8
Endrin aldehyde	6.169	6.073	6.273	45.950	50.000	-8.1
Endrin ketone	6.896	6.800	7.000	52.660	50.000	5.3
gamma-BHC (Lindane)	3.660	3.561	3.761	46.790	50.000	-6.4
gamma-Chlordane	5.042	4.945	5.145	46.450	50.000	-7.1
Heptachlor	4.007	3.909	4.109	45.300	50.000	-9.4
Heptachlor epoxide	4.790	4.693	4.893	47.030	50.000	-5.9
Methoxychlor	6.670	6.574	6.774	41.190	50.000	-17.6
Tetrachloro-m-xylene	2.824	2.724	2.924	42.200	50.000	-15.6