

Surrogate Summary

SDG No.: **Q3266**

Client: **CDM Smith**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PL097268.D	PIBLK-PL097268.D	Decachlorobiphen	1	20	17.3	87		57	171
		Tetrachloro-m-xyl	1	20	15.6	78		61	148
		Decachlorobiphen	2	20	17.1	86		57	171
		Tetrachloro-m-xyl	2	20	16.0	80		61	148
I.BLK-PL097493.D	PIBLK-PL097493.D	Decachlorobiphen	1	20	25.7	128		57	171
		Tetrachloro-m-xyl	1	20	21.2	106		61	148
		Decachlorobiphen	2	20	21.5	107		57	171
		Tetrachloro-m-xyl	2	20	22.4	112		61	148
PB169952BL	PB169952BL	Decachlorobiphen	1	20	25.5	128		20	144
		Tetrachloro-m-xyl	1	20	21.7	109		19	148
		Decachlorobiphen	2	20	21.6	108		20	144
		Tetrachloro-m-xyl	2	20	19.8	99		19	148
PB169952BS	PB169952BS	Decachlorobiphen	1	20	23.9	119		20	144
		Tetrachloro-m-xyl	1	20	20.3	102		19	148
		Decachlorobiphen	2	20	21.0	105		20	144
		Tetrachloro-m-xyl	2	20	19.1	96		19	148
Q3266-01	ENV-3-6FT	Decachlorobiphen	1	20	16.0	80		20	144
		Tetrachloro-m-xyl	1	20	11.7	58		19	148
		Decachlorobiphen	2	20	8.08	40		20	144
		Tetrachloro-m-xyl	2	20	11.1	55		19	148
Q3266-02	ENV-4-6FT	Decachlorobiphen	1	20	11.7	58		20	144
		Tetrachloro-m-xyl	1	20	12.3	61		19	148
		Decachlorobiphen	2	20	9.81	49		20	144
		Tetrachloro-m-xyl	2	20	11.7	58		19	148
Q3266-01MS	ENV-3-6FTMS	Decachlorobiphen	1	20	9.59	48		20	144
		Tetrachloro-m-xyl	1	20	11.9	59		19	148
		Decachlorobiphen	2	20	8.93	45		20	144
		Tetrachloro-m-xyl	2	20	12.0	60		19	148
Q3266-01MSD	ENV-3-6FTMSD	Decachlorobiphen	1	20	10.5	52		20	144
		Tetrachloro-m-xyl	1	20	12.6	63		19	148
		Decachlorobiphen	2	20	9.08	45		20	144
		Tetrachloro-m-xyl	2	20	11.6	58		19	148
I.BLK-PL097505.D	PIBLK-PL097505.D	Decachlorobiphen	1	20	24.9	124		57	171
		Tetrachloro-m-xyl	1	20	21.2	106		61	148
		Decachlorobiphen	2	20	21.5	107		57	171
		Tetrachloro-m-xyl	2	20	20.1	101		61	148