

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100625\
 Data File : BP025924.D
 Acq On : 06 Oct 2025 15:34
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 06 15:57:56 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 16:45:04 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	133	-0.02
2	1,4-Dioxane	40.000	37.753	5.6	120	0.00
3	Pyridine	40.000	40.331	-0.8	123	0.00
4	n-Nitrosodimethylamine	40.000	33.208	17.0	104	0.00
5 S	2-Fluorophenol	80.000	81.243	-1.6	123	0.00
6	Aniline	40.000	37.420	6.4	112	-0.02
7 S	Phenol-d6	80.000	79.169	1.0	117	0.00
8	2-Chlorophenol	40.000	41.478	-3.7	125	-0.02
9	Benzaldehyde	40.000	49.828	-24.6	187	-0.02
10 C	Phenol	40.000	40.539	-1.3	120	0.00
11	bis(2-Chloroethyl)ether	40.000	40.065	-0.2	120	-0.01
12	1,3-Dichlorobenzene	40.000	39.012	2.5	120	-0.02
13 C	1,4-Dichlorobenzene	40.000	38.348	4.1	119	-0.02
14	1,2-Dichlorobenzene	40.000	38.176	4.6	118	-0.02
15	Benzyl Alcohol	40.000	37.700	5.7	113	-0.02
16	2,2'-oxybis(1-Chloropropane	40.000	35.684	10.8	109	-0.04
17	2-Methylphenol	40.000	39.998	0.0	118	-0.01
18	Hexachloroethane	40.000	38.973	2.6	121	-0.03
19 P	n-Nitroso-di-n-propylamine	40.000	37.718	5.7	109	-0.02
20	3+4-Methylphenols	40.000	38.877	2.8	117	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	125	-0.02
22	Acetophenone	40.000	40.067	-0.2	113	-0.02
23 S	Nitrobenzene-d5	80.000	71.019	11.2	101	-0.02
24	Nitrobenzene	40.000	40.173	-0.4	113	-0.02
25	Isophorone	40.000	38.301	4.2	111	-0.02
26 C	2-Nitrophenol	40.000	44.479	-11.2	124	-0.02
27	2,4-Dimethylphenol	40.000	36.991	7.5	103	-0.02
28	bis(2-Chloroethoxy)methane	40.000	39.925	0.2	115	-0.03
29 C	2,4-Dichlorophenol	40.000	42.274	-5.7	119	0.00
30	1,2,4-Trichlorobenzene	40.000	39.061	2.3	115	-0.02
31	Naphthalene	40.000	39.632	0.9	116	-0.03
32	Benzoic acid	40.000	39.252	1.9	110	0.00
33	4-Chloroaniline	40.000	36.372	9.1	100	-0.02
34 C	Hexachlorobutadiene	40.000	41.789	-4.5	123	-0.04
35	Caprolactam	40.000	37.762	5.6	109	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.741	0.6	113	-0.01
37	2-Methylnaphthalene	40.000	43.959	-9.9	127	-0.03
38	1-Methylnaphthalene	40.000	39.718	0.7	114	-0.03
39 I	Acenaphthene-d10	20.000	20.000	0.0	121	-0.03
40	1,2,4,5-Tetrachlorobenzene	40.000	42.318	-5.8	117	-0.03
41 P	Hexachlorocyclopentadiene	40.000	37.735	5.7	99	-0.04
42 S	2,4,6-Tribromophenol	80.000	85.052	-6.3	116	-0.04
43 C	2,4,6-Trichlorophenol	40.000	42.952	-7.4	116	-0.03
44	2,4,5-Trichlorophenol	40.000	41.163	-2.9	110	0.00
45 S	2-Fluorobiphenyl	80.000	71.321	10.8	101	-0.03
46	1,1'-Biphenyl	40.000	40.199	-0.5	112	-0.03
47	2-Chloronaphthalene	40.000	40.278	-0.7	113	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100625\
 Data File : BP025924.D
 Acq On : 06 Oct 2025 15:34
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 06 15:57:56 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 16:45:04 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.768	0.6	106	-0.02
49	Acenaphthylene	40.000	35.723	10.7	100	-0.04
50	Dimethylphthalate	40.000	39.230	1.9	109	-0.04
51	2,6-Dinitrotoluene	40.000	39.077	2.3	107	-0.03
52 C	Acenaphthene	40.000	40.102	-0.3	112	-0.04
53	3-Nitroaniline	40.000	40.312	-0.8	105	-0.02
54 P	2,4-Dinitrophenol	40.000	37.020	7.4	104	0.00
55	Dibenzofuran	40.000	39.189	2.0	111	-0.04
56 P	4-Nitrophenol	40.000	39.940	0.2	108	0.02
57	2,4-Dinitrotoluene	40.000	40.790	-2.0	109	-0.02
58	Fluorene	40.000	39.313	1.7	110	-0.05
59	2,3,4,6-Tetrachlorophenol	40.000	39.559	1.1	107	-0.04
60	Diethylphthalate	40.000	39.006	2.5	111	-0.04
61	4-Chlorophenyl-phenylether	40.000	40.497	-1.2	115	-0.04
62	4-Nitroaniline	40.000	39.588	1.0	103	-0.03
63	Azobenzene	40.000	36.831	7.9	101	-0.04
64 I	Phenanthrene-d10	20.000	20.000	0.0	117	-0.03
65	4,6-Dinitro-2-methylphenol	40.000	43.634	-9.1	113	-0.02
66 c	n-Nitrosodiphenylamine	40.000	39.619	1.0	107	-0.03
67	4-Bromophenyl-phenylether	40.000	42.780	-7.0	116	-0.04
68	Hexachlorobenzene	40.000	42.973	-7.4	119	-0.03
69	Atrazine	40.000	39.198	2.0	104	-0.02
70 C	Pentachlorophenol	40.000	39.145	2.1	104	-0.02
71	Phenanthrene	40.000	39.434	1.4	108	-0.02
72	Anthracene	40.000	39.175	2.1	107	-0.02
73	Carbazole	40.000	39.864	0.3	107	-0.02
74	Di-n-butylphthalate	40.000	40.965	-2.4	110	-0.03
75 C	Fluoranthene	40.000	39.631	0.9	109	-0.02
76 I	Chrysene-d12	20.000	20.000	0.0	115	-0.03
77	Benzidine	40.000	43.710	-9.3	111	-0.02
78	Pyrene	40.000	41.048	-2.6	110	-0.02
79 S	Terphenyl-d14	80.000	72.461	9.4	98	-0.03
80	Butylbenzylphthalate	40.000	40.237	-0.6	104	-0.04
81	Benzo(a)anthracene	40.000	39.666	0.8	107	-0.03
82	3,3'-Dichlorobenzidine	40.000	38.336	4.2	100	-0.04
83	Chrysene	40.000	39.107	2.2	106	-0.03
84	Bis(2-ethylhexyl)phthalate	40.000	39.541	1.1	103	-0.05
85 c	Di-n-octyl phthalate	40.000	40.413	-1.0	106	-0.06
86 I	Perylene-d12	20.000	20.000	0.0	116	-0.06
87	Indeno(1,2,3-cd)pyrene	40.000	40.417	-1.0	107	-0.06
88	Benzo(b)fluoranthene	40.000	40.117	-0.3	107	-0.04
89	Benzo(k)fluoranthene	40.000	39.831	0.4	108	-0.04
90 C	Benzo(a)pyrene	40.000	38.667	3.3	103	-0.05
91	Dibenzo(a,h)anthracene	40.000	40.450	-1.1	108	-0.08
92	Benzo(g,h,i)perylene	40.000	39.324	1.7	106	-0.06

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100625\
Data File : BP025924.D
Acq On : 06 Oct 2025 15:34
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Oct 06 15:57:56 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 11 16:45:04 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0