

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3266

Analytical Method: SW8270E

Client: CDM Smith

DataFile: BP025936.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q3289-05MSD		Client Sample ID: 72-11966MSD									
Benzaldehyde	1100	0	720	ug/Kg	65		0		10	171	20
Phenol	1100	0	980	ug/Kg	89		1		51	122	20
bis(2-Chloroethyl)ether	1100	0	930	ug/Kg	85		0		54	125	20
2-Chlorophenol	1100	0	1000	ug/Kg	91		1		51	121	20
2-Methylphenol	1100	0	980	ug/Kg	89		2		47	125	20
2,2-oxybis(1-Chloropropane)	1100	0	830	ug/Kg	75		0		46	119	20
Acetophenone	1100	0	1100	ug/Kg	100		0		55	128	20
3+4-Methylphenols	1100	0	930	ug/Kg	85		1		49	125	20
N-Nitroso-di-n-propylamine	1100	0	890	ug/Kg	81		4		59	119	20
Hexachloroethane	1100	0	950	ug/Kg	86		1		51	116	20
Nitrobenzene	1100	0	960	ug/Kg	87		1		47	124	20
Isophorone	1100	0	920	ug/Kg	84		1		49	127	20
2-Nitrophenol	1100	0	1100	ug/Kg	100		9		43	131	20
2,4-Dimethylphenol	1100	0	1000	ug/Kg	91		0		63	151	20
bis(2-Chloroethoxy)methane	1100	0	950	ug/Kg	86		0		51	119	20
2,4-Dichlorophenol	1100	0	1100	ug/Kg	100		9		50	122	20
Naphthalene	1100	0	960	ug/Kg	87		0		51	121	20
4-Chloroaniline	1100	0	660	ug/Kg	60		5		10	100	20
Hexachlorobutadiene	1100	0	1000	ug/Kg	91		0		44	126	20
Caprolactam	1100	0	930	ug/Kg	85		5		51	134	20
4-Chloro-3-methylphenol	1100	0	1000	ug/Kg	91		4		57	132	20
2-Methylnaphthalene	1100	0	990	ug/Kg	90		5		59	123	20
Hexachlorocyclopentadiene	2200	0	1600	ug/Kg	73		0		10	175	20
2,4,6-Trichlorophenol	1100	0	1100	ug/Kg	100		0		33	141	20
2,4,5-Trichlorophenol	1100	0	1100	ug/Kg	100		9		38	135	20
1,1-Biphenyl	1100	0	1100	ug/Kg	100		0		55	131	20
2-Chloronaphthalene	1100	0	1000	ug/Kg	91		0		48	124	20
2-Nitroaniline	1100	0	980	ug/Kg	89		2		47	134	20
Dimethylphthalate	1100	0	950	ug/Kg	86		0		54	120	20
Acenaphthylene	1100	0	990	ug/Kg	90		1		57	125	20
2,6-Dinitrotoluene	1100	0	1000	ug/Kg	91		1		48	127	20
3-Nitroaniline	1100	0	730	ug/Kg	66		2		10	112	20
Acenaphthene	1100	0	960	ug/Kg	87		1		70	121	20
2,4-Dinitrophenol	2200	0	1800	ug/Kg	82		6		10	155	20
4-Nitrophenol	2200	0	1600	ug/Kg	73		7		10	175	20
Dibenzofuran	1100	0	960	ug/Kg	87		0		52	114	20
2,4-Dinitrotoluene	1100	0	1000	ug/Kg	91		0		41	140	20
Diethylphthalate	1100	0	970	ug/Kg	88		1		51	119	20
4-Chlorophenyl-phenylether	1100	0	1000	ug/Kg	91		0		48	122	20
Fluorene	1100	0	990	ug/Kg	90		1		53	118	20
4-Nitroaniline	1100	0	1000	ug/Kg	91		12		29	140	20
4,6-Dinitro-2-methylphenol	1100	0	1100	ug/Kg	100		9		10	160	20
N-Nitrosodiphenylamine	1100	0	1000	ug/Kg	91		0		73	118	20
4-Bromophenyl-phenylether	1100	0	1100	ug/Kg	100		9		65	121	20
Hexachlorobenzene	1100	0	1100	ug/Kg	100		0		67	118	20
Atrazine	1100	0	1100	ug/Kg	100		9		45	175	20

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3266

Analytical Method: SW8270E

Client: CDM Smith

DataFile: BP025936.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Pentachlorophenol	2200	0	2100	ug/Kg	95		0		13	153	20
Phenanthrene	1100	0	1000	ug/Kg	91		4		52	128	20
Anthracene	1100	0	1000	ug/Kg	91		2		62	124	20
Carbazole	1100	0	1000	ug/Kg	91		1		59	119	20
Di-n-butylphthalate	1100	0	1000	ug/Kg	91		1		55	125	20
Fluoranthene	1100	0	1000	ug/Kg	91		0		44	125	20
Pyrene	1100	0	1000	ug/Kg	91		1		37	122	20
Butylbenzylphthalate	1100	0	1000	ug/Kg	91		1		44	135	20
3,3-Dichlorobenzidine	1100	0	810	ug/Kg	74		6		15	112	20
Benzo(a)anthracene	1100	0	1000	ug/Kg	91		1		53	119	20
Chrysene	1100	0	1000	ug/Kg	91		1		57	121	20
bis(2-Ethylhexyl)phthalate	1100	0	1000	ug/Kg	91		2		42	169	20
Di-n-octyl phthalate	1100	0	1100	ug/Kg	100		9		51	156	20
Benzo(b)fluoranthene	1100	0	1000	ug/Kg	91		0		52	117	20
Benzo(k)fluoranthene	1100	0	1000	ug/Kg	91		0		57	134	20
Benzo(a)pyrene	1100	0	1000	ug/Kg	91		0		70	142	20
Indeno(1,2,3-cd)pyrene	1100	0	1100	ug/Kg	100		9		40	129	20
Dibenz(a,h)anthracene	1100	0	1100	ug/Kg	100		9		43	123	20
Benzo(g,h,i)perylene	1100	0	1100	ug/Kg	100		9		24	125	20
1,2,4,5-Tetrachlorobenzene	1100	0	1100	ug/Kg	100		9		52	134	20
1,4-Dioxane	1100	0	810	ug/Kg	74		0		46	112	20
2,3,4,6-Tetrachlorophenol	1100	0	1100	ug/Kg	100		9		24	146	20