

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3266

Client ID : SSTDCCC040

Date Analyzed: 10/02/2025

Lab File ID: BF143853.D

Time Analyzed: 13:36

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	114791	6.887	408348	8.17	201182	9.92
UPPER LIMIT	229582	7.387	816696	8.669	402364	10.422
LOWER LIMIT	57395.5	6.387	204174	7.669	100591	9.422
EPA SAMPLE NO.						
01 ENV-3-6FT	126293	6.88	481036	8.16	258523	9.92
02 ENV-3-6FTMS	119401	6.89	452831	8.17	241644	9.92
03 ENV-3-6FTMSD	132113	6.89	498372	8.17	264381	9.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.