

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3266

Client ID : SSTDCCC040

Date Analyzed: 10/08/2025

Lab File ID: BF143887.D

Time Analyzed: 10:13

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	125960	6.881	486402	8.17	269911	9.92
UPPER LIMIT	251920	7.381	972804	8.669	539822	10.422
LOWER LIMIT	62980	6.381	243201	7.669	134956	9.422
EPA SAMPLE NO.						
01 PB169953BL	127094	6.88	497763	8.16	273707	9.92
02 PB169953BS	137733	6.88	539635	8.16	299662	9.93
03 ENV-4-6FT	95054	6.88	368458	8.16	204631	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.