

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3266

Client ID : SSTDCCC040

Date Analyzed: 10/06/2025

Lab File ID: BP025924.D

Time Analyzed: 15:34

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	192886	7.737	769676	10.50	505298	14.35
UPPER LIMIT	385772	8.237	1539350	11.001	1010600	14.848
LOWER LIMIT	96443	7.237	384838	10.001	252649	13.848
EPA SAMPLE NO.						
01 72-11966MS	161138	7.74	617249	10.50	391056	14.34
02 72-11966MSD	153847	7.74	597298	10.50	396048	14.34
03 PB169989BL	166902	7.74	621599	10.50	409905	14.36
04 PB169989BS	182029	7.74	705808	10.50	446200	14.35

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.