

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CAMP02
 Lab Code: ACE SDG NO.: Q3266
 Lab File ID: VW032322.D Date Analyzed: 10/02/2025
 Instrument ID: MSVOA_W Time Analyzed: 10:53
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	303983	7.96	550095	8.85	496155	11.63
UPPER LIMIT	607966	8.459	1100190	9.349	992310	12.129
LOWER LIMIT	151992	7.459	275048	8.349	248078	11.129
EPA SAMPLE NO.						
ENV-3-6FT	180169	7.97	409718	8.86	401656	11.63
ENV-4-6FT	184689	7.97	407370	8.86	435664	11.63
VW1002SBL01	200843	7.97	433970	8.85	450698	11.63
VW1002SBS01	277189	7.97	526280	8.85	484324	11.63
VW1002SBSD01	291487	7.97	534672	8.86	467867	11.63

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.