

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q3267

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB137435

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.095	0.099	96	85-115	10/06/2025
Sample ID: Cyanide	CCV1	mg/L	0.24	0.25	96	90-110	10/06/2025
Sample ID: Cyanide	CCV2	mg/L	0.24	0.25	96	90-110	10/06/2025

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q3267

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB137498

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Field pH		pH	7.01	7	100	90-110	10/01/2025
Sample ID:	CCV1						
Field pH		pH	7.00	7	100	90-110	10/01/2025
Sample ID:	CCV2						
Field pH		pH	7.01	7	100	90-110	10/01/2025

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q3267

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB137498

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------