

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: METEM A GE POWER Business

SDG No.: Q3267

Contract: METE01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	1.10	1.0	110	-5	7	10/08/2025	12:45	LB137466
	Chromium	47.7	52.0	92	42	62	10/08/2025	12:45	LB137466
	Copper	16.3	2.0	815	-18	22	10/08/2025	12:45	LB137466
	Lead	-4.56			-12	12	10/08/2025	12:45	LB137466
	Nickel	3.02	2.0	151	-38	42	10/08/2025	12:45	LB137466
	Silver	7.27			-10	10	10/08/2025	12:45	LB137466
	Zinc	1.82			-40	40	10/08/2025	12:45	LB137466
ICSAB01	Cadmium	1010	972	104	826	1120	10/08/2025	12:50	LB137466
	Chromium	552	542	102	460	624	10/08/2025	12:50	LB137466
	Copper	496	511	97	434	588	10/08/2025	12:50	LB137466
	Lead	41.6	49.0	85	37	61	10/08/2025	12:50	LB137466
	Nickel	1000	954	105	810	1100	10/08/2025	12:50	LB137466
	Silver	219	201	109	170	232	10/08/2025	12:50	LB137466
	Zinc	1050	952	110	809	1095	10/08/2025	12:50	LB137466
ICSA	Cadmium	-1.84	1.0	184	-5	7	10/08/2025	12:54	LB137466
	Chromium	52.9	52.0	102	42	62	10/08/2025	12:54	LB137466
	Copper	24.1	2.0	1205	-18	22	10/08/2025	12:54	LB137466
	Lead	-13.4			-12	12	10/08/2025	12:54	LB137466
	Nickel	3.26	2.0	163	-38	42	10/08/2025	12:54	LB137466
	Silver	6.92			-10	10	10/08/2025	12:54	LB137466
	Zinc	1.80			-40	40	10/08/2025	12:54	LB137466
ICSAB	Cadmium	957	972	98	826	1120	10/08/2025	12:59	LB137466
	Chromium	588	542	108	460	624	10/08/2025	12:59	LB137466
	Copper	538	511	105	434	588	10/08/2025	12:59	LB137466
	Lead	45.5	49.0	93	37	61	10/08/2025	12:59	LB137466
	Nickel	972	954	102	810	1100	10/08/2025	12:59	LB137466
	Silver	208	201	104	170	232	10/08/2025	12:59	LB137466
	Zinc	1060	952	111	809	1095	10/08/2025	12:59	LB137466