

661374

Test results

Aquakem 7.2AQ1

Page:

Alliance Technical Group  
284 Sheffield Street, Mountainside, NJ 07092Reviewed by : RM

Instrument ID : Konelab

10/3/2025 11:15

Test: Ammonia-N

| Sample Id     | Result | Dil. 1 + | Response | Errors          |
|---------------|--------|----------|----------|-----------------|
| ICV1          | 0.982  | 0.0      | 0.215    |                 |
| ICB1          | 0.009  | 0.0      | 0.026    |                 |
| CCV1          | 0.970  | 0.0      | 0.212    |                 |
| CCB1          | 0.009  | 0.0      | 0.026    |                 |
| RL CHECK      | 0.093  | 0.0      | 0.043    |                 |
| PB169946BL    | 0.009  | 0.0      | 0.026    |                 |
| PB169946BS    | 1.027  | 0.0      | 0.223    |                 |
| Q3254-01      | 0.019  | 0.0      | 0.028    |                 |
| Q3254-03      | 11.838 | 0.0      | 2.317    | Test limit high |
| Q3254-05      | 4.803  | 0.0      | 0.955    | Test limit high |
| Q3254-07      | 11.716 | 0.0      | 2.293    | Test limit high |
| Q3258-06      | 0.316  | 0.0      | 0.086    |                 |
| Q3263-01      | 0.016  | 0.0      | 0.028    |                 |
| Q3263-02      | 0.010  | 0.0      | 0.027    |                 |
| CCV2          | 0.968  | 0.0      | 0.212    |                 |
| CCB2          | 0.011  | 0.0      | 0.027    |                 |
| Q3263-02DUP   | 0.008  | 0.0      | 0.026    |                 |
| Q3263-02MS    | 1.038  | 0.0      | 0.226    |                 |
| Q3263-02MSD   | 1.041  | 0.0      | 0.226    |                 |
| Q3268-01      | 0.639  | 0.0      | 0.148    |                 |
| CCV3          | 1.014  | 0.0      | 0.221    |                 |
| CCB3          | 0.015  | 0.0      | 0.028    |                 |
| Q3254-03DLX10 | 1.107  | 0.0      | 0.239    |                 |
| Q3254-05DLX5  | 0.919  | 0.0      | 0.202    |                 |
| Q3254-07DLX10 | 1.118  | 0.0      | 0.241    |                 |
| CCV4          | 0.978  | 0.0      | 0.214    |                 |
| CCB4          | 0.011  | 0.0      | 0.027    |                 |
| N             | 27     |          |          |                 |
| Mean          | 1.507  |          |          |                 |
| SD            | 3.1074 |          |          |                 |
| CV%           | 206.22 |          |          |                 |

93% CSO-ISO  
10/03/2025  
RM

Aquakem v. 7.2AQ1

Results from time period:

Fri Oct 03 09:28:04 2025

Fri Oct 03 11:10:19 2025

| Sample Id     | Sam/Ctr/c/ | Test short r | Test type | Result  | Result unit | Result date and time | Stat |
|---------------|------------|--------------|-----------|---------|-------------|----------------------|------|
| 0.0PPM        | A          | Ammonia-† P  |           | 0.0099  | mg/l        | 10/3/2025 9:28:04    |      |
| 0.1PPM        | A          | Ammonia-† P  |           | 0.1173  | mg/l        | 10/3/2025 9:28:05    |      |
| 0.2PPM        | A          | Ammonia-† P  |           | 0.2087  | mg/l        | 10/3/2025 9:28:06    |      |
| 0.4PPM        | A          | Ammonia-† P  |           | 0.3836  | mg/l        | 10/3/2025 9:28:07    |      |
| 1.0PPM        | A          | Ammonia-† P  |           | 0.9767  | mg/l        | 10/3/2025 9:28:08    |      |
| 1.3PPM        | A          | Ammonia-† P  |           | 1.3053  | mg/l        | 10/3/2025 9:28:09    |      |
| 2.0PPM        | A          | Ammonia-† P  |           | 2.0319  | mg/l        | 10/3/2025 9:28:10    |      |
| ICV1          | S          | Ammonia-† P  |           | 0.9818  | mg/l        | 10/3/2025 10:18:24   |      |
| ICB1          | S          | Ammonia-† P  |           | 0.0094  | mg/l        | 10/3/2025 10:18:26   |      |
| CCV1          | S          | Ammonia-† P  |           | 0.9696  | mg/l        | 10/3/2025 10:18:29   |      |
| CCB1          | S          | Ammonia-† P  |           | 0.0089  | mg/l        | 10/3/2025 10:18:31   |      |
| RL CHECK      | S          | Ammonia-† P  |           | 0.0932  | mg/l        | 10/3/2025 10:18:33   |      |
| PB169946BL    | S          | Ammonia-† P  |           | 0.0095  | mg/l        | 10/3/2025 10:29:09   |      |
| PB169946BS    | S          | Ammonia-† P  |           | 1.0265  | mg/l        | 10/3/2025 10:29:12   |      |
| Q3254-01      | S          | Ammonia-† P  |           | 0.0192  | mg/l        | 10/3/2025 10:29:13   |      |
| Q3254-03      | S          | Ammonia-† P  |           | 11.8381 | mg/l        | 10/3/2025 10:29:14   |      |
| Q3254-05      | S          | Ammonia-† P  |           | 4.8029  | mg/l        | 10/3/2025 10:29:15   |      |
| Q3254-07      | S          | Ammonia-† P  |           | 11.716  | mg/l        | 10/3/2025 10:29:16   |      |
| Q3258-06      | S          | Ammonia-† P  |           | 0.316   | mg/l        | 10/3/2025 10:29:17   |      |
| Q3263-01      | S          | Ammonia-† P  |           | 0.0156  | mg/l        | 10/3/2025 10:29:18   |      |
| Q3263-02      | S          | Ammonia-† P  |           | 0.0103  | mg/l        | 10/3/2025 10:29:19   |      |
| CCV2          | S          | Ammonia-† P  |           | 0.9675  | mg/l        | 10/3/2025 10:39:53   |      |
| CCB2          | S          | Ammonia-† P  |           | 0.0108  | mg/l        | 10/3/2025 10:39:56   |      |
| Q3263-02DUP   | S          | Ammonia-† P  |           | 0.0078  | mg/l        | 10/3/2025 10:39:57   |      |
| Q3263-02MS    | S          | Ammonia-† P  |           | 1.0379  | mg/l        | 10/3/2025 10:39:59   |      |
| Q3263-02MSD   | S          | Ammonia-† P  |           | 1.0414  | mg/l        | 10/3/2025 10:40:00   |      |
| Q3268-01      | S          | Ammonia-† P  |           | 0.6394  | mg/l        | 10/3/2025 10:40:01   |      |
| CCV3          | S          | Ammonia-† P  |           | 1.0144  | mg/l        | 10/3/2025 10:46:30   |      |
| CCB3          | S          | Ammonia-† P  |           | 0.0151  | mg/l        | 10/3/2025 10:46:32   |      |
| Q3254-03DLX10 | S          | Ammonia-† P  |           | 1.107   | mg/l        | 10/3/2025 11:10:10   |      |
| Q3254-05DLX5  | S          | Ammonia-† P  |           | 0.9186  | mg/l        | 10/3/2025 11:10:13   |      |
| Q3254-07DLX10 | S          | Ammonia-† P  |           | 1.1184  | mg/l        | 10/3/2025 11:10:14   |      |
| CCV4          | S          | Ammonia-† P  |           | 0.9777  | mg/l        | 10/3/2025 11:10:16   |      |
| CCB4          | S          | Ammonia-† P  |           | 0.0106  | mg/l        | 10/3/2025 11:10:18   |      |

Calibration results

Aquakem 7.2AQ1

Page:

Alliance Technical Group

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

10/3/2025 9:30

Test Ammonia-N


Accepted 10/3/2025 9:30

Factor 5.164

Bias 0.025

Coeff. of det. 0.999084

Errors



|   | Calibrator | Response | Calc. con. | Conc.  | Errors |
|---|------------|----------|------------|--------|--------|
| 1 | 0.00PPM    | 0.026    | 0.0099     | 0.0000 | -      |
| 2 | NH3-2PPM   | 0.047    | 0.1173     | 0.1000 | 17.3   |
| 3 | NH3-2PPM   | 0.065    | 0.2087     | 0.2000 | 4.3    |
| 4 | NH3-2PPM   | 0.099    | 0.3836     | 0.4000 | -4.1   |
| 5 | NH3-2PPM   | 0.214    | 0.9767     | 1.0000 | -2.3   |
| 6 | NH3-2PPM   | 0.277    | 1.3053     | 1.3333 | 0.4    |
| 7 | NH3-2PPM   | 0.418    | 2.0319     | 2.0000 | 1.6    |

10/03/2025  
RM