284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	**SDG No.:**	Q3268
Project:	Rotor Clip NJ WTD - 2025	**Run No.:**	LB137395

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID LB137395BS								
Residual Chlorine	**mg/L**	**0.4**	**0.41**		**102**	**1**	**90-110**	**10/02/2025**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	**SDG No.:**	Q3268
Project:	Rotor Clip NJ WTD - 2025	**Run No.:**	LB137417

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID PB169946BS								
Ammonia as N	**mg/L**	**1**	**1.00**		**100**	**1**	**90-110**	**10/03/2025**