

SOP ID : M3010A-Digestion-17

SDG No : N/A

Matrix : WATER

Pipette ID: ICP A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M6069

Hood ID : #3

Block ID: 1. HOT BOCK #1 2. N/A

Start Digest Date: 10/03/2025 Time : 10:20 Temp : 96 °C

End Digest Date: 10/03/2025 Time : 13:25 Temp : 96 °C

Digestion tube ID: M5595

Block thermometer ID: MET-DIG. #1

Dig Technician Signature: SKJ.

Supervisor Signature:

Temp : 1. 96°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	0.25	M6180
LFS-2	0.25	M6181
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Conc. HNO3	3.00	M6158
1:1 HCL	5.00	MP87148
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

HOT BLOCK#1 CELL#50 96 C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
10/03/2025 13:25	SKJ met dig	SKJ (metals lab)
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
PB169978BL	PBW978	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	1
PB169978BS	LCS978	<2	50	25	Colorless	Colorless	Clear	Clear	M6180,M6181	2
Q3268-01MS	WATER-TREATMENT-DISCHA RCEMS	<2	50	25	Colorless	Colorless	Clear	Clear	M6180,M6181	5
Q3268-01MSD	WATER-TREATMENT-DISCHA RCEMSD	<2	50	25	Colorless	Colorless	Clear	Clear	M6180,M6181	6
Q3268-01DUP	WATER-TREATMENT-DISCHA RCE DUP	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	4
Q3268-01	WATER-TREATMENT-DISCHA RGE	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	3
Q3275-01	MW3	<2	50	25	light Brown	Colorless	Clear	Clear	N/A	7
Q3275-02	MW4	<2	50	25	Brown	light Brown	Cloudy	Clear	N/A	8
Q3275-03	MW5	<2	50	25	light Brown	Colorless	Clear	Clear	N/A	9
Q3275-04	MW1	<2	50	25	light Brown	Colorless	Clear	Clear	N/A	10
Q3278-01	MW10	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	11