



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3268 **Order ID:** Q3268
Client: VERINA CONSULTING GROUP, LLC **Project ID:** Rotor Clip NJ WTD - 2025

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : WATER-TREATMENT-DISCHARGE								
Q3268-01	WATER-TREATMENT-DISCH	Water	Chromium	114		1.06	5.00	ug/L
Q3268-01	WATER-TREATMENT-DISCH	Water	Copper	28.4		2.30	10.0	ug/L
Q3268-01	WATER-TREATMENT-DISCH	Water	Nickel	21.4		1.53	20.0	ug/L
Q3268-01	WATER-TREATMENT-DISCH	Water	Zinc	3310		8.33	20.0	ug/L



SAMPLE DATA

Report of Analysis

Client:	VERINA CONSULTING GROUP, LLC	Date Collected:	10/02/25
Project:	Rotor Clip NJ WTD - 2025	Date Received:	10/02/25
Client Sample ID:	WATER-TREATMENT-DISCHARGE	SDG No.:	Q3268
Lab Sample ID:	Q3268-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-47-3	Chromium	114		1	1.06	5.00	ug/L	10/03/25 10:20	10/08/25 14:43	6010D	SW3010
7440-50-8	Copper	28.4		1	2.30	10.0	ug/L	10/03/25 10:20	10/08/25 14:43	6010D	SW3010
7440-02-0	Nickel	21.4		1	1.53	20.0	ug/L	10/03/25 10:20	10/08/25 14:43	6010D	SW3010
7440-66-6	Zinc	3310		1	8.33	20.0	ug/L	10/03/25 10:20	10/08/25 14:43	6010D	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3268

Contract: VERI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Chromium	2.12	+/-5	U	10.0	P	10/08/2025	12:36	LB137466
	Copper	4.60	+/-10	U	20.0	P	10/08/2025	12:36	LB137466
	Nickel	3.06	+/-20	U	40.0	P	10/08/2025	12:36	LB137466
	Zinc	16.7	+/-20	U	40.0	P	10/08/2025	12:36	LB137466

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3268

Contract: VERI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Chromium	2.12	+/-5	U	10.0	P	10/08/2025	13:16	LB137466
	Copper	4.60	+/-10	U	20.0	P	10/08/2025	13:16	LB137466
	Nickel	3.06	+/-20	U	40.0	P	10/08/2025	13:16	LB137466
	Zinc	16.7	+/-20	U	40.0	P	10/08/2025	13:16	LB137466
CCB02	Chromium	2.12	+/-5	U	10.0	P	10/08/2025	14:18	LB137466
	Copper	4.60	+/-10	U	20.0	P	10/08/2025	14:18	LB137466
	Nickel	3.06	+/-20	U	40.0	P	10/08/2025	14:18	LB137466
	Zinc	16.7	+/-20	U	40.0	P	10/08/2025	14:18	LB137466
CCB03	Chromium	2.12	+/-5	U	10.0	P	10/08/2025	15:13	LB137466
	Copper	4.60	+/-10	U	20.0	P	10/08/2025	15:13	LB137466
	Nickel	3.06	+/-20	U	40.0	P	10/08/2025	15:13	LB137466
	Zinc	16.7	+/-20	U	40.0	P	10/08/2025	15:13	LB137466
CCB04	Chromium	2.12	+/-5	U	10.0	P	10/08/2025	16:15	LB137466
	Copper	4.60	+/-10	U	20.0	P	10/08/2025	16:15	LB137466
	Nickel	3.06	+/-20	U	40.0	P	10/08/2025	16:15	LB137466
	Zinc	16.7	+/-20	U	40.0	P	10/08/2025	16:15	LB137466
CCB05	Chromium	2.12	+/-5	U	10.0	P	10/08/2025	17:06	LB137466
	Copper	4.60	+/-10	U	20.0	P	10/08/2025	17:06	LB137466
	Nickel	3.06	+/-20	U	40.0	P	10/08/2025	17:06	LB137466
	Zinc	16.7	+/-20	U	40.0	P	10/08/2025	17:06	LB137466
CCB06	Chromium	2.12	+/-5	U	10.0	P	10/08/2025	17:57	LB137466
	Copper	4.60	+/-10	U	20.0	P	10/08/2025	17:57	LB137466
	Nickel	3.06	+/-20	U	40.0	P	10/08/2025	17:57	LB137466
	Zinc	16.7	+/-20	U	40.0	P	10/08/2025	17:57	LB137466

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3268

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169978BL	WATER			Batch Number:	PB169978		Prep Date:	10/03/2025	
	Chromium	1.06	<2.5	U	5.00	P	10/08/2025	15:30	LB137466
	Copper	2.30	<5	U	10.0	P	10/08/2025	15:30	LB137466
	Nickel	1.53	<10	U	20.0	P	10/08/2025	15:30	LB137466
	Zinc	8.33	<10	U	20.0	P	10/08/2025	15:30	LB137466



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3268

Contract: VERI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Chromium	793	800	99	90 - 110	P	10/08/2025	12:24	LB137466
	Copper	999	1000	100	90 - 110	P	10/08/2025	12:24	LB137466
	Nickel	1950	2000	98	90 - 110	P	10/08/2025	12:24	LB137466
	Zinc	1990	2000	99	90 - 110	P	10/08/2025	12:24	LB137466

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3268

Contract: VERI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Chromium	9.92	10.0	99	80 - 120	P	10/08/2025	12:32	LB137466
	Copper	21.7	20.0	108	80 - 120	P	10/08/2025	12:32	LB137466
	Nickel	39.6	40.0	99	80 - 120	P	10/08/2025	12:32	LB137466
	Zinc	42.7	40.0	107	80 - 120	P	10/08/2025	12:32	LB137466

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3268

Contract: VERI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Chromium	989	1000	99	90 - 110	P	10/08/2025	13:12	LB137466
	Copper	1240	1250	100	90 - 110	P	10/08/2025	13:12	LB137466
	Nickel	2470	2500	99	90 - 110	P	10/08/2025	13:12	LB137466
	Zinc	2480	2500	99	90 - 110	P	10/08/2025	13:12	LB137466
CCV02	Chromium	951	1000	95	90 - 110	P	10/08/2025	14:14	LB137466
	Copper	1190	1250	95	90 - 110	P	10/08/2025	14:14	LB137466
	Nickel	2360	2500	94	90 - 110	P	10/08/2025	14:14	LB137466
	Zinc	2330	2500	93	90 - 110	P	10/08/2025	14:14	LB137466
CCV03	Chromium	943	1000	94	90 - 110	P	10/08/2025	15:04	LB137466
	Copper	1170	1250	94	90 - 110	P	10/08/2025	15:04	LB137466
	Nickel	2320	2500	93	90 - 110	P	10/08/2025	15:04	LB137466
	Zinc	2380	2500	95	90 - 110	P	10/08/2025	15:04	LB137466
CCV04	Chromium	965	1000	96	90 - 110	P	10/08/2025	16:11	LB137466
	Copper	1190	1250	95	90 - 110	P	10/08/2025	16:11	LB137466
	Nickel	2340	2500	94	90 - 110	P	10/08/2025	16:11	LB137466
	Zinc	2370	2500	95	90 - 110	P	10/08/2025	16:11	LB137466
CCV05	Chromium	942	1000	94	90 - 110	P	10/08/2025	17:02	LB137466
	Copper	1180	1250	94	90 - 110	P	10/08/2025	17:02	LB137466
	Nickel	2310	2500	92	90 - 110	P	10/08/2025	17:02	LB137466
	Zinc	2460	2500	98	90 - 110	P	10/08/2025	17:02	LB137466
CCV06	Chromium	949	1000	95	90 - 110	P	10/08/2025	17:53	LB137466
	Copper	1200	1250	96	90 - 110	P	10/08/2025	17:53	LB137466
	Nickel	2350	2500	94	90 - 110	P	10/08/2025	17:53	LB137466
	Zinc	2410	2500	96	90 - 110	P	10/08/2025	17:53	LB137466



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3268

Contract: VERI01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Chromium	9.98	10.0	100	65 - 135	P	10/08/2025	12:41	LB137466
	Copper	21.6	20.0	108	65 - 135	P	10/08/2025	12:41	LB137466
	Nickel	39.3	40.0	98	65 - 135	P	10/08/2025	12:41	LB137466
	Zinc	43.1	40.0	108	65 - 135	P	10/08/2025	12:41	LB137466

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>VERINA CONSULTING GROUP, LLC</u>	SDG No.: <u>Q3268</u>
Contract: <u>VERI01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Chromium	47.7	52.0	92	42	62	10/08/2025	12:45	LB137466
	Copper	16.3	2.0	815	-18	22	10/08/2025	12:45	LB137466
	Nickel	3.02	2.0	151	-38	42	10/08/2025	12:45	LB137466
	Zinc	1.82			-40	40	10/08/2025	12:45	LB137466
ICSAB01	Chromium	552	542	102	460	624	10/08/2025	12:50	LB137466
	Copper	496	511	97	434	588	10/08/2025	12:50	LB137466
	Nickel	1000	954	105	810	1100	10/08/2025	12:50	LB137466
	Zinc	1050	952	110	809	1095	10/08/2025	12:50	LB137466



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>VERINA CONSULTING GROUP, LLC</u>	level: <u>low</u>	sdg no.: <u>Q3268</u>
contract: <u>VERI01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3268-01</u>	client id: <u>WATER-TREATMENT-DISCHARGEMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3268-01MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	75 - 125	299		114		200	93		P
Copper	ug/L	75 - 125	162		28.4		150	89		P
Nickel	ug/L	75 - 125	257		21.4		250	94		P
Zinc	ug/L	75 - 125	3430		3310		100	114		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>VERINA CONSULTING GROUP, LLC</u>	level: <u>low</u>	sdg no.: <u>Q3268</u>
contract: <u>VERI01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3268-01</u>	client id: <u>WATER-TREATMENT-DISCHARGEMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3268-01MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	75 - 125	301		114		200	94		P
Copper	ug/L	75 - 125	162		28.4		150	89		P
Nickel	ug/L	75 - 125	257		21.4		250	94		P
Zinc	ug/L	75 - 125	3470		3310		100	155		P

Metals
- 5b -

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3268

Contract: VERI01

Lab Code: ACE

Matrix: _____

Level: LOW

Client ID: _____

Sample ID: _____ **Spiked ID:** _____

Analyte	Units	Acceptance Limit %R	C	Sample Result	C	Spike Added	% Recovery	Qual	M
---------	-------	------------------------	---	------------------	---	----------------	---------------	------	---

.....

A
B
C
D
E
F
G
H
I
J

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** Q3268
Contract: VERI01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q3268-01 **Client ID:** WATER-TREATMENT-DISCHARGEDU
Percent Solids for Sample: NA **Duplicate ID** Q3268-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Chromium	ug/L	20	114		116		2		P
Copper	ug/L	20	28.4		28.7		1		P
Nickel	ug/L	20	21.4		21.6		1		P
Zinc	ug/L	20	3310		3460		4		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** Q3268
Contract: VERI01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q3268-01MS **Client ID:** WATER-TREATMENT-DISCHARGE
Percent Solids for Sample: NA **Duplicate ID** Q3268-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Chromium	ug/L	20	299		301		1		P
Copper	ug/L	20	162		162		0		P
Nickel	ug/L	20	257		257		0		P
Zinc	ug/L	20	3430		3470		1		P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3268

Contract: VERI01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169978BS							
Chromium	ug/L	200	212		106	80 - 120	P
Copper	ug/L	150	154		103	80 - 120	P
Nickel	ug/L	250	247		99	80 - 120	P
Zinc	ug/L	100	100		100	80 - 120	P

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

WATER-TREATMENT-DISCHARGEL

Lab Name: Alliance **Contract:** VERI01
Lab Code: ACE **Lb No.:** lb137466 **Lab Sample ID :** Q3268-01L **SDG No.:** Q3268
Matrix (soil/water): Water **Level (low/med):** LOW
Concentration Units: ug/L

Analyte	Initial Sample Result (I) C		Serial Dilution Result (S) C		% Differ- ence		Q	M
Chromium	114		120		6			P
Copper	28.4		33.5	J	18			P
Nickel	21.4		22.7	J	6			P
Zinc	3310		3530		7			P

metals
- 14 -
ANALYSIS RUN LOG

Client: VERINA CONSULTING GROUP, LLC

Contract: VERI01

Lab code: ACE

Sdg no.: Q3268

Instrument id number:

Method:

Run number: LB137466

Start date: 10/08/2025

End date: 10/08/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1200	Cr,Cu,Ni,Zn
S1	S1	1	1204	Cr,Cu,Ni,Zn
S2	S2	1	1208	Cr,Cu,Ni,Zn
S3	S3	1	1212	Cr,Cu,Ni,Zn
S4	S4	1	1216	Cr,Cu,Ni,Zn
S5	S5	1	1220	Cr,Cu,Ni,Zn
ICV01	ICV01	1	1224	Cr,Cu,Ni,Zn
LLICV01	LLICV01	1	1232	Cr,Cu,Ni,Zn
ICB01	ICB01	1	1236	Cr,Cu,Ni,Zn
CRI01	CRI01	1	1241	Cr,Cu,Ni,Zn
ICSA01	ICSA01	1	1245	Cr,Cu,Ni,Zn
ICSAB01	ICSAB01	1	1250	Cr,Cu,Ni,Zn
CCV01	CCV01	1	1312	Cr,Cu,Ni,Zn
CCB01	CCB01	1	1316	Cr,Cu,Ni,Zn
CCV02	CCV02	1	1414	Cr,Cu,Ni,Zn
CCB02	CCB02	1	1418	Cr,Cu,Ni,Zn
Q3268-01	WATER-TREATMENT-DISCH/	1	1443	Cr,Cu,Ni,Zn
Q3268-01DUP	WATER-TREATMENT-DISCH/	1	1447	Cr,Cu,Ni,Zn
Q3268-01L	WATER-TREATMENT-DISCH/	5	1452	Cr,Cu,Ni,Zn
Q3268-01MS	WATER-TREATMENT-DISCH/	1	1456	Cr,Cu,Ni,Zn
Q3268-01MSD	WATER-TREATMENT-DISCH/	1	1500	Cr,Cu,Ni,Zn
CCV03	CCV03	1	1504	Cr,Cu,Ni,Zn
CCB03	CCB03	1	1513	Cr,Cu,Ni,Zn
PB169978BL	PB169978BL	1	1530	Cr,Cu,Ni,Zn
PB169978BS	PB169978BS	1	1534	Cr,Cu,Ni,Zn
CCV04	CCV04	1	1611	Cr,Cu,Ni,Zn
CCB04	CCB04	1	1615	Cr,Cu,Ni,Zn
CCV05	CCV05	1	1702	Cr,Cu,Ni,Zn
CCB05	CCB05	1	1706	Cr,Cu,Ni,Zn
CCV06	CCV06	1	1753	Cr,Cu,Ni,Zn
CCB06	CCB06	1	1757	Cr,Cu,Ni,Zn



METAL PREPARATION & INSTRUMENT DATA

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3268

Contract: VERI01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0007850	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0001050	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3268

Contract: VERI01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		As	Ba	Be	Cd	Co
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0000000	0.0009530
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3268

Contract: VERI01

Lab Code: ACE

Instrument ID:

Date:

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Cr	Cu	K	Mn	Mo
Chromium	267.716	0.0000000	0.0000000	0.0000070	0.0002200	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0006510	0.0020500
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0009010	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3268

Contract: VERI01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	-0.0047000	0.0036100	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0067600	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3268

Contract: VERI01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0001110	0.0000000
Copper	224.700	0.0000000	0.0003840	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000



LAB CHRONICLE

OrderID:	Q3268	OrderDate:	10/2/2025 12:46:00 PM
Client:	VERINA CONSULTING GROUP, LLC	Project:	Rotor Clip NJ WTD - 2025
Contact:	Michael Valenzi	Location:	D31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3268-01	WATER-TREATMENT-D ISCHARGE	Water	Metals Group4	6010D	10/02/25	10/03/25	10/08/25	10/02/25



METAL PREPARATION & ANALYICAL SUMMARY

Metals
- 13 -

SAMPLE PREPARATION SUMMARY

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q3268
Contract: VERI01 **Lab Code:** ACE **Method:** _____

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB169978							
PB169978BL	PB169978BL	MB	WATER	10/03/2025	50.0	25.0	
PB169978BS	PB169978BS	LCS	WATER	10/03/2025	50.0	25.0	
Q3268-01	WATER-TREATMENT-DISCHARGE	SAM	WATER	10/03/2025	50.0	25.0	
Q3268-01DUP	WATER-TREATMENT-DISCHARGEDUP	DUP	WATER	10/03/2025	50.0	25.0	
Q3268-01MS	WATER-TREATMENT-DISCHARGEMS	MS	WATER	10/03/2025	50.0	25.0	
Q3268-01MSD	WATER-TREATMENT-DISCHARGEMSD	MSD	WATER	10/03/2025	50.0	25.0	

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137466

Review By	Janvi	Review On	10/10/2025 1:31:26 PM
Supervise By	jaswal	Supervise On	10/10/2025 10:40:09 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032		
ICV Standard	MP87047,MP87038		
CCV Standard	MP87042		
ICSA Standard	MP87040,MP87048		
CRI Standard	MP87038		
LCS Standard			
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/08/25 12:00		Jaswal	OK
2	S1	S1	CAL2	10/08/25 12:04		Jaswal	OK
3	S2	S2	CAL3	10/08/25 12:08		Jaswal	OK
4	S3	S3	CAL4	10/08/25 12:12		Jaswal	OK
5	S4	S4	CAL5	10/08/25 12:16		Jaswal	OK
6	S5	S5	CAL6	10/08/25 12:20		Jaswal	OK
7	ICV01	ICV01	ICV	10/08/25 12:24		Jaswal	OK
8	LLICV01	LLICV01	LLICV	10/08/25 12:32		Jaswal	OK
9	ICB01	ICB01	ICB	10/08/25 12:36		Jaswal	OK
10	CRI01	CRI01	CRDL	10/08/25 12:41		Jaswal	OK
11	ICSA01	ICSA01	ICSA	10/08/25 12:45		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	10/08/25 12:50		Jaswal	OK
13	ICSADL	ICSADL	ICSA	10/08/25 12:54		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	10/08/25 12:59		Jaswal	OK
15	CCV01	CCV01	CCV	10/08/25 13:12		Jaswal	OK
16	CCB01	CCB01	CCB	10/08/25 13:16		Jaswal	OK
17	LR-2	LR-2	HIGH STD	10/08/25 13:25		Jaswal	OK
18	LR-3	LR-3	HIGH STD	10/08/25 13:29		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137466

Review By	Janvi	Review On	10/10/2025 1:31:26 PM
Supervise By	jaswal	Supervise On	10/10/2025 10:40:09 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032		
ICV Standard	MP87047,MP87038		
CCV Standard	MP87042		
ICSA Standard	MP87040,MP87048		
CRI Standard	MP87038		
LCS Standard			
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049		

19	Q3295-03	EO-02-10062025	SAM	10/08/25 13:34		Jaswal	OK
20	Q3295-01	EO-01-10062025	SAM	10/08/25 13:38		Jaswal	OK
21	Q3295-01L	EO-01-10062025L	SD	10/08/25 13:46		Jaswal	OK
22	Q3295-01MS	EO-01-10062025MS	MS	10/08/25 13:50		Jaswal	OK
23	Q3295-01MSD	EO-01-10062025MSD	MSD	10/08/25 13:54		Jaswal	OK
24	Q3295-01A	EO-01-10062025A	PS	10/08/25 13:58	0.1ML M6008,M6017-10ML SAMPLE	Jaswal	OK
25	LR-1	LR-1	HIGH STD	10/08/25 14:02		Jaswal	OK
26	Q3295-01DUP	EO-01-10062025DUP	DUP	10/08/25 14:08		Jaswal	OK
27	CCV02	CCV02	CCV	10/08/25 14:14		Jaswal	OK
28	CCB02	CCB02	CCB	10/08/25 14:18		Jaswal	OK
29	Q3296-01	SU-04-10062025	SAM	10/08/25 14:23		Jaswal	OK
30	Q3293-01	CONC-STORAGE-PA	SAM	10/08/25 14:27		Jaswal	OK
31	Q3294-01	HD-02-10062025	SAM	10/08/25 14:31		Jaswal	OK
32	Q3297-01	TP1	SAM	10/08/25 14:35		Jaswal	OK
33	Q3297-11	TP2	SAM	10/08/25 14:39		Jaswal	OK
34	Q3268-01	WATER-TREATMENT	SAM	10/08/25 14:43		Jaswal	OK
35	Q3268-01DUP	WATER-TREATMENT	DUP	10/08/25 14:47		Jaswal	OK
36	Q3268-01L	WATER-TREATMENT	SD	10/08/25 14:52		Jaswal	OK
37	Q3268-01MS	WATER-TREATMENT	MS	10/08/25 14:56		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137466

Review By	Janvi	Review On	10/10/2025 1:31:26 PM
Supervise By	jaswal	Supervise On	10/10/2025 10:40:09 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032		
ICV Standard	MP87047,MP87038		
CCV Standard	MP87042		
ICSA Standard	MP87040,MP87048		
CRI Standard	MP87038		
LCS Standard			
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049		

38	Q3268-01MSD	WATER-TREATMENT	MSD	10/08/25 15:00		Jaswal	OK
39	CCV03	CCV03	CCV	10/08/25 15:04		Jaswal	OK
40	CCB03	CCB03	CCB	10/08/25 15:13		Jaswal	OK
41	Q3268-01A	WATER-TREATMENT	PS	10/08/25 15:17	0.1ML M6008,M6017-10ML SAMPLE	Jaswal	OK
42	PB169978BL	PB169978BL	MB	10/08/25 15:30		Jaswal	OK
43	PB169978BS	PB169978BS	LCS	10/08/25 15:34		Jaswal	OK
44	PB169959BL	PB169959BL	MB	10/08/25 15:38		Jaswal	OK
45	PB169959BS	PB169959BS	LCS	10/08/25 15:43		Jaswal	OK
46	Q3275-01	MW3	SAM	10/08/25 15:47		Jaswal	OK
47	Q3275-02	MW4	SAM	10/08/25 15:51		Jaswal	OK
48	Q3275-03	MW5	SAM	10/08/25 15:55	Na High	Jaswal	Dilution
49	CCV04	CCV04	CCV	10/08/25 16:11		Jaswal	OK
50	CCB04	CCB04	CCB	10/08/25 16:15		Jaswal	OK
51	Q3275-04	MW1	SAM	10/08/25 16:19	Not Use	Jaswal	Not Ok
52	Q3278-01	MW10	SAM	10/08/25 16:24	Not Use	Jaswal	Not Ok
53	Q3260-02	COMP	SAM	10/08/25 16:28		Jaswal	OK
54	Q3260-02DUP	COMPDUP	DUP	10/08/25 16:32		Jaswal	OK
55	Q3260-02L	COMPL	SD	10/08/25 16:37		Jaswal	OK
56	Q3260-02MS	COMPMS	MS	10/08/25 16:41		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137466

Review By	Janvi	Review On	10/10/2025 1:31:26 PM
Supervise By	jaswal	Supervise On	10/10/2025 10:40:09 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032		
ICV Standard	MP87047,MP87038		
CCV Standard	MP87042		
ICSA Standard	MP87040,MP87048		
CRI Standard	MP87038		
LCS Standard			
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049		

57	Q3260-02MSD	COMPMSD	MSD	10/08/25 16:45		Jaswal	OK
58	Q3260-02A	COMPA	PS	10/08/25 16:49	0.1ML M6008,M6017-10ML SAMPLE	Jaswal	OK
59	Q3267-01DL2	Q4DL2	SAM	10/08/25 16:53	Not Use	Jaswal	Not Ok
60	Q3267-01DL	Q4DL	SAM	10/08/25 16:58	50X Straight Dilution	Jaswal	OK
61	CCV05	CCV05	CCV	10/08/25 17:02		Jaswal	OK
62	CCB05	CCB05	CCB	10/08/25 17:06		Jaswal	OK
63	PB170024BL	PB170024BL	MB	10/08/25 17:10		Jaswal	OK
64	PB170024BS	PB170024BS	LCS	10/08/25 17:15		Jaswal	OK
65	Q3293-02	CONC-STORAGE-PA	SAM	10/08/25 17:19		Jaswal	OK
66	Q3297-10	TP1	SAM	10/08/25 17:23		Jaswal	OK
67	Q3297-20	TP2	SAM	10/08/25 17:27		Jaswal	OK
68	Q3302-06	OR-658-COMP-01	SAM	10/08/25 17:32	NOT USE	Jaswal	Not Ok
69	Q3302-06DUP	OR-658-COMP-01DU	DUP	10/08/25 17:36	NOT USE	Jaswal	Not Ok
70	Q3302-06L	OR-658-COMP-01L	SD	10/08/25 17:41	NOT USE	Jaswal	Not Ok
71	Q3302-06MS	OR-658-COMP-01MS	MS	10/08/25 17:45	NOT USE	Jaswal	Not Ok
72	Q3302-06MSD	OR-658-COMP-01MS	MSD	10/08/25 17:49	NOT USE	Jaswal	Not Ok
73	CCV06	CCV06	CCV	10/08/25 17:53		Jaswal	OK
74	CCB06	CCB06	CCB	10/08/25 17:57		Jaswal	OK

SOP ID : M3010A-Digestion-17

SDG No : N/A

Matrix : WATER

Pipette ID: ICP A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M6069

Hood ID : #3

Block ID: 1. HOT BOCK #1 2. N/A

Start Digest Date: 10/03/2025 Time : 10:20 Temp : 96 °C

End Digest Date: 10/03/2025 Time : 13:25 Temp : 96 °C

Digestion tube ID: M5595

Block thermometer ID: MET-DIG. #1

Dig Technician Signature: SKJ

Supervisor Signature: [Signature]

Temp : 1. 96°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	0.25	M6180
LFS-2	0.25	M6181
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Conc. HNO3	3.00	M6158
1:1 HCL	5.00	MP87148
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

HOT BLOCK#1 CELL#50 96 C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
10/03/2025 13:25	SKJ met dig	SKJ (metals lab)
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
PB169978BL	PBW978	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	1
PB169978BS	LCS978	<2	50	25	Colorless	Colorless	Clear	Clear	M6180,M6181	2
Q3268-01MS	WATER-TREATMENT-DISCHA RCEMS	<2	50	25	Colorless	Colorless	Clear	Clear	M6180,M6181	5
Q3268-01MSD	WATER-TREATMENT-DISCHA RCEMSD	<2	50	25	Colorless	Colorless	Clear	Clear	M6180,M6181	6
Q3268-01DUP	WATER-TREATMENT-DISCHA RCE DUP	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	4
Q3268-01	WATER-TREATMENT-DISCHA RCE	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	3
Q3275-01	MW3	<2	50	25	light Brown	Colorless	Clear	Clear	N/A	7
Q3275-02	MW4	<2	50	25	Brown	light Brown	Cloudy	Clear	N/A	8
Q3275-03	MW5	<2	50	25	light Brown	Colorless	Clear	Clear	N/A	9
Q3275-04	MW1	<2	50	25	light Brown	Colorless	Clear	Clear	N/A	10
Q3278-01	MW10	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	11