

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3268

Contract: VERI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Chromium	793	800	99	90 - 110	P	10/08/2025	12:24	LB137466
	Copper	999	1000	100	90 - 110	P	10/08/2025	12:24	LB137466
	Nickel	1950	2000	98	90 - 110	P	10/08/2025	12:24	LB137466
	Zinc	1990	2000	99	90 - 110	P	10/08/2025	12:24	LB137466

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3268

Contract: VERI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Chromium	9.92	10.0	99	80 - 120	P	10/08/2025	12:32	LB137466
	Copper	21.7	20.0	108	80 - 120	P	10/08/2025	12:32	LB137466
	Nickel	39.6	40.0	99	80 - 120	P	10/08/2025	12:32	LB137466
	Zinc	42.7	40.0	107	80 - 120	P	10/08/2025	12:32	LB137466

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3268

Contract: VERI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Chromium	989	1000	99	90 - 110	P	10/08/2025	13:12	LB137466
	Copper	1240	1250	100	90 - 110	P	10/08/2025	13:12	LB137466
	Nickel	2470	2500	99	90 - 110	P	10/08/2025	13:12	LB137466
	Zinc	2480	2500	99	90 - 110	P	10/08/2025	13:12	LB137466
CCV02	Chromium	951	1000	95	90 - 110	P	10/08/2025	14:14	LB137466
	Copper	1190	1250	95	90 - 110	P	10/08/2025	14:14	LB137466
	Nickel	2360	2500	94	90 - 110	P	10/08/2025	14:14	LB137466
	Zinc	2330	2500	93	90 - 110	P	10/08/2025	14:14	LB137466
CCV03	Chromium	943	1000	94	90 - 110	P	10/08/2025	15:04	LB137466
	Copper	1170	1250	94	90 - 110	P	10/08/2025	15:04	LB137466
	Nickel	2320	2500	93	90 - 110	P	10/08/2025	15:04	LB137466
	Zinc	2380	2500	95	90 - 110	P	10/08/2025	15:04	LB137466
CCV04	Chromium	965	1000	96	90 - 110	P	10/08/2025	16:11	LB137466
	Copper	1190	1250	95	90 - 110	P	10/08/2025	16:11	LB137466
	Nickel	2340	2500	94	90 - 110	P	10/08/2025	16:11	LB137466
	Zinc	2370	2500	95	90 - 110	P	10/08/2025	16:11	LB137466
CCV05	Chromium	942	1000	94	90 - 110	P	10/08/2025	17:02	LB137466
	Copper	1180	1250	94	90 - 110	P	10/08/2025	17:02	LB137466
	Nickel	2310	2500	92	90 - 110	P	10/08/2025	17:02	LB137466
	Zinc	2460	2500	98	90 - 110	P	10/08/2025	17:02	LB137466
CCV06	Chromium	949	1000	95	90 - 110	P	10/08/2025	17:53	LB137466
	Copper	1200	1250	96	90 - 110	P	10/08/2025	17:53	LB137466
	Nickel	2350	2500	94	90 - 110	P	10/08/2025	17:53	LB137466
	Zinc	2410	2500	96	90 - 110	P	10/08/2025	17:53	LB137466