

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3268

Contract: VERI01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Chromium	47.7	52.0	92	42	62	10/08/2025	12:45	LB137466
	Copper	16.3	2.0	815	-18	22	10/08/2025	12:45	LB137466
	Nickel	3.02	2.0	151	-38	42	10/08/2025	12:45	LB137466
	Zinc	1.82			-40	40	10/08/2025	12:45	LB137466
ICSAB01	Chromium	552	542	102	460	624	10/08/2025	12:50	LB137466
	Copper	496	511	97	434	588	10/08/2025	12:50	LB137466
	Nickel	1000	954	105	810	1100	10/08/2025	12:50	LB137466
	Zinc	1050	952	110	809	1095	10/08/2025	12:50	LB137466