

metals
- 5a -
MATRIX SPIKE SUMMARY

client: VERINA CONSULTING GROUP, LLC **level:** low **sdg no.:** Q3268
contract: VERI01 **lab code:** ACE
matrix: Water **sample id:** Q3268-01 **client id:** WATER-TREATMENT-DISCHARGEMS
Percent Solids for Sample: NA **Spiked ID:** Q3268-01MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	75 - 125	299		114		200	93		P
Copper	ug/L	75 - 125	162		28.4		150	89		P
Nickel	ug/L	75 - 125	257		21.4		250	94		P
Zinc	ug/L	75 - 125	3430		3310		100	114		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>VERINA CONSULTING GROUP, LLC</u>	level: <u>low</u>	sdg no.: <u>Q3268</u>
contract: <u>VERI01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3268-01</u>	client id: <u>WATER-TREATMENT-DISCHARGEMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3268-01MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	75 - 125	301		114		200	94		P
Copper	ug/L	75 - 125	162		28.4		150	89		P
Nickel	ug/L	75 - 125	257		21.4		250	94		P
Zinc	ug/L	75 - 125	3470		3310		100	155		P