

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** Q3268
Contract: VERI01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q3268-01 **Client ID:** WATER-TREATMENT-DISCHARGEDU
Percent Solids for Sample: NA **Duplicate ID** Q3268-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Chromium	ug/L	20	114		116		2		P
Copper	ug/L	20	28.4		28.7		1		P
Nickel	ug/L	20	21.4		21.6		1		P
Zinc	ug/L	20	3310		3460		4		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** Q3268
Contract: VERI01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q3268-01MS **Client ID:** WATER-TREATMENT-DISCHARGE
Percent Solids for Sample: NA **Duplicate ID** Q3268-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Chromium	ug/L	20	299		301		1		P
Copper	ug/L	20	162		162		0		P
Nickel	ug/L	20	257		257		0		P
Zinc	ug/L	20	3430		3470		1		P