

**Metals**

- 7 -

**LABORATORY CONTROL SAMPLE SUMMARY**

**Client:** VERINA CONSULTING GROUP, LLC

**SDG No.:** Q3268

**Contract:** VERI01

**Lab Code:** ACE

| Analyte    | Units | True Value | Result | C | %<br>Recovery | Acceptance<br>Limits | M |
|------------|-------|------------|--------|---|---------------|----------------------|---|
| PB169978BS |       |            |        |   |               |                      |   |
| Chromium   | ug/L  | 200        | 212    |   | 106           | 80 - 120             | P |
| Copper     | ug/L  | 150        | 154    |   | 103           | 80 - 120             | P |
| Nickel     | ug/L  | 250        | 247    |   | 99            | 80 - 120             | P |
| Zinc       | ug/L  | 100        | 100    |   | 100           | 80 - 120             | P |