

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG No.: Q3273

Instrument ID: BNA_G

Calibration Date(s): 10/02/2025 10/02/2025

Calibration Time(s): 09:01 15:49

LAB FILE ID:		RRF2.5 = BG064473.D			RRF005 = BG064474.D		RRF020 = BG064476.D	
		RRF040 = BG064477.D			RRF050 = BG064478.D		RRF060 = BG064479.D	
COMPOUND	RRF2.5	RRF005	RRF020	RRF040	RRF050	RRF060	RRF	% RSD
2-Fluorophenol		1.066	1.053	1.195	1.071	1.108	1.081	5.3
Benzaldehyde			1.039	1.527	1.152	1.548	1.290	18.7
Phenol-d6		1.395	1.446	1.631	1.434	1.496	1.452	6.1
Phenol		1.518	1.530	1.683	1.517	1.583	1.532	5.7
bis(2-Chloroethyl) ether		1.117	1.002	1.134	1.000	1.064	1.038	6.6
2-Chlorophenol		1.387	1.312	1.455	1.264	1.336	1.327	5.5
2-Methylphenol		1.105	1.087	1.215	1.076	1.143	1.109	4.8
2,2-oxybis(1-Chloropropane)		1.720	1.657	1.804	1.650	1.708	1.672	4.8
Acetophenone		0.476	0.477	0.516	0.491	0.495	0.484	4.2
3+4-Methylphenols			1.510	1.700	1.492	1.622	1.548	5.9
n-Nitroso-di-n-propylamine	0.813	1.008	1.006	1.096	0.959	1.048	0.971	9.1
Nitrobenzene-d5		0.360	0.378	0.396	0.369	0.381	0.372	3.8
Hexachloroethane		0.586	0.534	0.604	0.562	0.566	0.575	5.9
Nitrobenzene		0.403	0.352	0.395	0.364	0.377	0.371	5.7
Isophorone		0.672	0.674	0.721	0.660	0.686	0.674	3.6
2-Nitrophenol		0.197	0.181	0.192	0.185	0.188	0.185	4.9
2,4-Dimethylphenol		0.277	0.279	0.299	0.285	0.298	0.282	5.0
bis(2-Chloroethoxy) methane		0.320	0.361	0.375	0.336	0.352	0.343	5.9
2,4-Dichlorophenol		0.323	0.321	0.343	0.322	0.336	0.326	3.8
Naphthalene		1.062	1.013	1.107	0.995	1.041	1.030	4.3
4-Chloroaniline		0.392	0.373	0.417	0.388	0.406	0.395	3.5
Hexachlorobutadiene		0.331	0.308	0.316	0.296	0.300	0.304	4.9
Caprolactam			0.120	0.115	0.111	0.110	0.112	4.4
4-Chloro-3-methylphenol		0.395	0.392	0.418	0.390	0.405	0.397	3.0
2-Methylnaphthalene		0.783	0.790	0.843	0.778	0.799	0.791	3.2
Hexachlorocyclopentadiene			0.296	0.363	0.356	0.378	0.338	12.0
2,4,6-Trichlorophenol		0.391	0.396	0.445	0.416	0.436	0.413	5.3
2-Fluorobiphenyl		1.355	1.346	1.434	1.325	1.389	1.357	3.1
2,4,5-Trichlorophenol		0.451	0.430	0.486	0.465	0.473	0.461	3.9
1,1-Biphenyl		1.439	1.417	1.521	1.434	1.484	1.438	3.4
2-Chloronaphthalene		1.006	1.046	1.153	1.056	1.106	1.074	4.6
2-Nitroaniline		0.301	0.371	0.397	0.386	0.385	0.367	8.6
Dimethylphthalate		1.564	1.582	1.624	1.531	1.586	1.566	2.7
Acenaphthylene		1.591	1.573	1.704	1.574	1.638	1.594	3.7

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG No.: Q3273

Instrument ID: BNA_G

Calibration Date(s): 10/02/2025 10/02/2025

Calibration Time(s): 09:01 15:49

LAB FILE ID:		RRF2.5 = BG064473.D			RRF005 = BG064474.D		RRF020 = BG064476.D	
		RRF040 = BG064477.D			RRF050 = BG064478.D		RRF060 = BG064479.D	
COMPOUND	RRF2.5	RRF005	RRF020	RRF040	RRF050	RRF060	RRF	% RSD
2,6-Dinitrotoluene		0.292	0.335	0.337	0.333	0.345	0.326	5.4
3-Nitroaniline		0.279	0.300	0.319	0.315	0.303	0.301	4.5
Acenaphthene		1.108	1.086	1.159	1.106	1.138	1.108	3.0
2,4-Dinitrophenol			0.270	0.207	0.228	0.229	0.237	10.8
4-Nitrophenol			0.227	0.260	0.256	0.253	0.237	10.8
Dibenzofuran		1.837	1.800	1.909	1.821	1.872	1.829	2.8
2,4-Dinitrotoluene		0.481	0.488	0.527	0.513	0.506	0.499	3.6
Diethylphthalate		1.747	1.679	1.730	1.675	1.663	1.680	3.2
4-Chlorophenyl-phenylether		0.866	0.839	0.868	0.817	0.819	0.830	3.4
Fluorene		1.518	1.535	1.601	1.488	1.532	1.520	2.9
4-Nitroaniline		0.229	0.331	0.334	0.345	0.340	0.314	13.0
4,6-Dinitro-2-methylphenol			0.139	0.145	0.138	0.145	0.139	5.6
n-Nitrosodiphenylamine		0.540	0.534	0.569	0.508	0.541	0.534	3.6
2,4,6-Tribromophenol		0.364	0.367	0.381	0.364	0.361	0.363	2.9
4-Bromophenyl-phenylether		0.235	0.227	0.253	0.224	0.237	0.235	4.2
Hexachlorobenzene		0.256	0.272	0.288	0.260	0.280	0.269	4.4
Atrazine		0.237	0.247	0.250	0.232	0.239	0.238	4.3
Pentachlorophenol			0.146	0.169	0.166	0.170	0.157	13.7
Phenanthrene		1.072	1.055	1.111	1.002	1.050	1.046	3.6
Anthracene		1.045	1.044	1.130	1.001	1.068	1.049	3.9
Carbazole		0.883	0.904	0.987	0.877	0.921	0.907	4.2
Di-n-butylphthalate		1.096	1.105	1.189	1.076	1.131	1.107	3.7
Fluoranthene		1.269	1.272	1.330	1.192	1.253	1.250	3.8
Pyrene		1.295	1.316	1.392	1.242	1.281	1.298	4.2
Terphenyl-d14		1.067	1.069	1.123	1.030	1.044	1.057	3.6
Butylbenzylphthalate		0.437	0.464	0.497	0.447	0.479	0.460	4.8
3,3-Dichlorobenzidine			0.431	0.458	0.423	0.443	0.432	3.6
Benzo (a) anthracene		1.280	1.285	1.353	1.238	1.291	1.277	3.1
Chrysene		1.226	1.195	1.297	1.173	1.210	1.211	3.5
Bis (2-ethylhexyl) phthalate		0.605	0.640	0.692	0.641	0.670	0.649	4.3
Di-n-octyl phthalate			1.050	1.153	1.073	1.125	1.091	3.7
Benzo (b) fluoranthene		1.077	1.171	1.212	1.142	1.169	1.138	4.4
Benzo (k) fluoranthene		1.139	1.145	1.233	1.123	1.156	1.146	4.0
Benzo (a) pyrene		1.017	1.059	1.122	1.031	1.075	1.049	3.9
Indeno (1,2,3-cd) pyrene		1.368	1.378	1.493	1.387	1.431	1.392	4.0

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG No.: Q3273

Instrument ID: BNA_G

Calibration Date(s): 10/02/2025 10/02/2025

Calibration Time(s): 09:01 15:49

LAB FILE ID:		RRF2.5 = BG064473.D		RRF005 = BG064474.D		RRF020 = BG064476.D		
		RRF040 = BG064477.D		RRF050 = BG064478.D		RRF060 = BG064479.D		
COMPOUND	RRF2.5	RRF005	RRF020	RRF040	RRF050	RRF060	RRF	% RSD
Dibenzo (a,h) anthracene		1.047	1.121	1.190	1.115	1.151	1.113	4.6
Benzo (g,h,i) perylene		1.059	1.088	1.153	1.069	1.104	1.082	3.4
1,2,4,5-Tetrachlorobenzene		0.622	0.622	0.708	0.640	0.678	0.646	5.6
1,4-Dioxane		0.530	0.447	0.438	0.398	0.386	0.427	12.1
2,3,4,6-Tetrachlorophenol		0.438	0.475	0.505	0.475	0.494	0.470	5.2