

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QC Batch ID # BG100225

Review By	Rahul	Review On	10/3/2025 8:38:44 AM
Supervise By	Jagrut	Supervise On	10/3/2025 3:25:38 PM
SubDirectory	BG100225	HP Acquire Method	BNA_G
		HP Processing Method	bg100225

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13176,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BG064472.D	2 Oct 2025 8:21		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BG064473.D	2 Oct 2025 9:01		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BG064474.D	2 Oct 2025 9:42		RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BG064475.D	2 Oct 2025 10:53	Not Used	RC/JU	Not Ok
5	SSTDICC020	SSTDICC020	BG064476.D	2 Oct 2025 11:33		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BG064477.D	2 Oct 2025 12:14	Compound failed in the calibration. #77	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BG064478.D	2 Oct 2025 12:55		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BG064479.D	2 Oct 2025 13:35		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BG064480.D	2 Oct 2025 14:16		RC/JU	Ok,M
10	SSTDICC010	SSTDICC010	BG064481.D	2 Oct 2025 15:49		RC/JU	Ok,M
11	SSTDICV040	ICVBG100225	BG064482.D	2 Oct 2025 18:09		RC/JU	Ok,M
12	PB169936BL	PB169936BL	BG064483.D	2 Oct 2025 18:50	Surrogate Fail	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QC Batch ID # BG100325

Review By	Rahul	Review On	10/6/2025 4:14:36 PM		
Supervise By	Jagrut	Supervise On	10/6/2025 4:23:45 PM		
SubDirectory	BG100325	HP Acquire Method	BNA_G	HP Processing Method	bg100225

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13177,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BG064484.D	3 Oct 2025 12:52		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BG064485.D	3 Oct 2025 13:33		RC/JU	Ok,M
3	PB169969BL	PB169969BL	BG064486.D	3 Oct 2025 14:13		RC/JU	Ok
4	PB169969BS	PB169969BS	BG064487.D	3 Oct 2025 14:54		RC/JU	Ok,M
5	PB169973BL	PB169973BL	BG064488.D	3 Oct 2025 15:35		RC/JU	Ok
6	PB169973BS	PB169973BS	BG064489.D	3 Oct 2025 16:16		RC/JU	Ok,M
7	PB169973BSD	PB169973BSD	BG064490.D	3 Oct 2025 16:57		RC/JU	Ok,M
8	Q3263-01	251818	BG064491.D	3 Oct 2025 17:38		RC/JU	Ok
9	Q3263-02	266380	BG064492.D	3 Oct 2025 18:19		RC/JU	Ok
10	Q3273-01	MW1	BG064493.D	3 Oct 2025 18:59		RC/JU	Ok
11	Q3274-01	MW4	BG064494.D	3 Oct 2025 19:40		RC/JU	Ok
12	Q3272-06	RT5417-CONCRETE	BG064495.D	3 Oct 2025 20:21		RC/JU	Ok

M : Manual Integration