

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BG064472.D
Instrument ID: BNA_G

Contract: GENV01
SDG NO.: Q3274
DFTPP Injection Date: 10/02/2025
DFTPP Injection Time: 08:21

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.3 (1.1) 1
69	Mass 69 relative abundance	29.9
70	Less than 2.0% of mass 69	0.1 (0.4) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.4
365	Greater than 1% of mass 198	6.7
441	Present, but less than mass 443	16.8
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.5 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BG064473.D	10/02/2025	09:01
SSTDICC005	SSTDICC005	BG064474.D	10/02/2025	09:42
SSTDICC020	SSTDICC020	BG064476.D	10/02/2025	11:33
SSTDICCC040	SSTDICCC040	BG064477.D	10/02/2025	12:14
SSTDICC050	SSTDICC050	BG064478.D	10/02/2025	12:55
SSTDICC060	SSTDICC060	BG064479.D	10/02/2025	13:35
SSTDICC080	SSTDICC080	BG064480.D	10/02/2025	14:16
SSTDICC010	SSTDICC010	BG064481.D	10/02/2025	15:49