

Surrogate Summary

SDG No.: **Q3274**

Client: **G Environmental**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3274-01	MW4	1,2-Dichloroethane-d4	50	52.3	105		70 (74)	130 (125)
		Dibromofluoromethane	50	45.1	90		70 (75)	130 (124)
		Toluene-d8	50	49.9	100		70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.9	106		70 (77)	130 (121)
Q3274-01DL	MW4DL	1,2-Dichloroethane-d4	50	53.8	108		70 (74)	130 (125)
		Dibromofluoromethane	50	47.3	95		70 (75)	130 (124)
		Toluene-d8	50	49.5	99		70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.7	101		70 (77)	130 (121)
Q3274-02	MW5	1,2-Dichloroethane-d4	50	49.4	99		70 (74)	130 (125)
		Dibromofluoromethane	50	43.3	87		70 (75)	130 (124)
		Toluene-d8	50	45.8	92		70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.1	98		70 (77)	130 (121)
Q3274-02DL	MW5DL	1,2-Dichloroethane-d4	50	52.1	104		70 (74)	130 (125)
		Dibromofluoromethane	50	47.1	94		70 (75)	130 (124)
		Toluene-d8	50	49.6	99		70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.7	101		70 (77)	130 (121)
VN1003WBL01	VN1003WBL01	1,2-Dichloroethane-d4	50	57.1	114		70 (74)	130 (125)
		Dibromofluoromethane	50	47.8	96		70 (75)	130 (124)
		Toluene-d8	50	47.2	94		70 (86)	130 (113)
		4-Bromofluorobenzene	50	42.0	84		70 (77)	130 (121)
VN1003WBS01	VN1003WBS01	1,2-Dichloroethane-d4	50	52.0	104		70 (74)	130 (125)
		Dibromofluoromethane	50	49.9	100		70 (75)	130 (124)
		Toluene-d8	50	49.8	100		70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.8	98		70 (77)	130 (121)
VN1003WBSD0	VN1003WBSD01	1,2-Dichloroethane-d4	50	47.4	95		70 (74)	130 (125)
		Dibromofluoromethane	50	43.1	86		70 (75)	130 (124)
		Toluene-d8	50	44.1	88		70 (86)	130 (113)
		4-Bromofluorobenzene	50	44.4	89		70 (77)	130 (121)

