

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
Lab Code: ACE SDG NO.: Q3274
Lab File ID: VN088013.D Date Analyzed: 10/03/2025
Instrument ID: MSVOA_N Time Analyzed: 10:17
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	211900	8.21	407793	9.08	399348	11.85
UPPER LIMIT	423800	8.706	815586	9.582	798696	12.347
LOWER LIMIT	105950	7.706	203897	8.582	199674	11.347
EPA SAMPLE NO.						
MW4	261013	8.21	577276	9.08	582521	11.84
MW4DL	193443	8.21	426224	9.08	419833	11.85
MW5	274294	8.21	584137	9.08	570238	11.84
MW5DL	199029	8.21	430919	9.08	412320	11.85
VN1003WBL01	175124	8.21	396641	9.08	390371	11.84
VN1003WBS01	210745	8.21	397174	9.08	382298	11.85
VN1003WBSD01	228703	8.21	434830	9.08	402581	11.85

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.