

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3274
 Lab File ID: VN088013.D Date Analyzed: 10/03/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:17
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	212700	13.77				
UPPER LIMIT	425400	14.27				
LOWER LIMIT	106350	13.27				
EPA SAMPLE NO.						
MW4	296966	13.77				
MW4DL	205320	13.77				
MW5	295091	13.77				
MW5DL	199129	13.77				
VN1003WBL01	175413	13.77				
VN1003WBS01	201044	13.77				
VN1003WBSD01	206765	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.