



284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

CASE NARRATIVE

G Environmental

Project Name: Banker

Project # N/A

Order ID # Q3275

Test Name: Metals Group4, Anions Group1

A. Number of Samples and Date of Receipt:

4 Water samples were received on 10/02/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Metals Group4, Anions Group1. This data package contains results for Metals Group4(6010D), Anions Group1(300.0).

C. Analytical Techniques:

Metals Group4 : The analysis of Metals Group4 was based on method 6010D and digestion based on method 3010 (waters).

Wetchem : The analysis of Anions Group1 was based on method 300.0.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

Metals Group4 : Sample MW5 was diluted due to high concentrations for Sodium.

Wetchem : Sample MW3 was diluted due to high concentrations for Chloride & Sample MW4 was diluted due to high concentrations for Chloride & Sample MW5 was diluted due to high concentrations for Chloride & Sample MW5DL was diluted due to high concentrations for Chloride & Sample MW1 was diluted due to high concentrations for Chloride.

The MS recoveries met the requirements for all compounds except following Wetchem : The Matrix Spike (MW1MS) analysis met criteria for all compounds except for Chloride due to sample matrix interference.

The MSD recoveries met the requirements for all compounds except following Wetchem : The Matrix Spike Duplicate (MW1MSD) analysis met criteria for all compounds except for Chloride due to sample matrix interference.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Duplicate analysis met criteria for all compound.

The Serial Dilution met the acceptable requirements.



284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

E. Additional Comments:

Metals Group4 : In analytical sequence LB137466, The Result was outside of acceptance limit for Sodium of CCB05 and CCB06 but, no any sample associated under these CCBs.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____